

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062322\
 Data File : VX029753.D
 Acq On : 23 Jun 2022 18:47
 Operator : JC/MD
 Sample : N3354-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP061422

Quant Time: Jun 24 05:41:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	250990	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	408993	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	362701	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	167831	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	135294	40.721	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	81.440%		
35) Dibromofluoromethane	5.391	113	122471	45.949	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.900%		
50) Toluene-d8	8.653	98	469216	48.189	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.380%		
62) 4-Bromofluorobenzene	11.085	95	176422	48.046	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.100%		
Target Compounds					Qvalue	
16) Acetone	2.380	43	30773	23.237	ug/l	97
17) Carbon Disulfide	2.532	76	1976	0.300	ug/l	100
25) 2-Butanone	4.568	43	17149	8.193	ug/l	90
31) Cyclohexane	5.470	56	31918	7.156	ug/l	95
39) Methylcyclohexane	7.379	83	19502	4.336	ug/l #	56
40) Benzene	6.044	78	147459	13.790	ug/l	100
52) Toluene	8.720	92	8083	1.180	ug/l	96
67) Ethyl Benzene	10.195	91	26955	2.105	ug/l	98
68) m/p-Xylenes	10.305	106	8928	1.788	ug/l	91
69) o-Xylene	10.646	106	2300	0.466	ug/l	90
73) Isopropylbenzene	10.963	105	279370	22.201	ug/l	98
78) n-propylbenzene	11.305	91	294207	20.759	ug/l	98
83) tert-Butylbenzene	11.719	119	5289	0.512	ug/l	82
85) sec-Butylbenzene	11.890	105	36095	2.819	ug/l #	56
89) n-Butylbenzene	12.335	91	17960	1.985	ug/l #	44

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062322\
Data File : VX029753.D
Acq On : 23 Jun 2022 18:47
Operator : JC/MD
Sample : N3354-23
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP061422

Quant Time: Jun 24 05:41:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

