

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062322\
 Data File : VX029755.D
 Acq On : 23 Jun 2022 19:34
 Operator : JC/MD
 Sample : N3365-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-21D

Quant Time: Jun 24 05:41:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	260310	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	434162	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	370012	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	168196	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	139720	40.548	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	81.100%
35) Dibromofluoromethane	5.391	113	125732	44.438	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.880%
50) Toluene-d8	8.653	98	489430	47.351	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.700%
62) 4-Bromofluorobenzene	11.079	95	175185	44.943	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.880%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	3944	2.872	ug/l	92
17) Carbon Disulfide	2.514	76	5682	0.831	ug/l	98
39) Methylcyclohexane	7.385	83	1253	0.262	ug/l	92
52) Toluene	8.726	92	1835	0.252	ug/l	88
67) Ethyl Benzene	10.195	91	4075	0.312	ug/l	91
84) 1,2,4-Trimethylbenzene	11.756	105	5370	0.507	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062322\
Data File : VX029755.D
Acq On : 23 Jun 2022 19:34
Operator : JC/MD
Sample : N3365-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-21D

Quant Time: Jun 24 05:41:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

