

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062323\
 Data File : VX036292.D
 Acq On : 23 Jun 2023 15:59
 Operator : JC/MD
 Sample : VX0623WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0623WBL01

Quant Time: Jun 23 16:24:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	220127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	391199	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	318619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136533	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	166852	52.879	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.760%	
35) Dibromofluoromethane	5.379	113	130773	50.698	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.400%	
50) Toluene-d8	8.647	98	426151	44.233	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 88.460%#	
62) 4-Bromofluorobenzene	11.079	95	149145	43.950	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 87.900%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062323\
Data File : VX036292.D
Acq On : 23 Jun 2023 15:59
Operator : JC/MD
Sample : VX0623WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0623WBL01

Quant Time: Jun 23 16:24:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

