

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062325\
 Data File : VX046809.D
 Acq On : 23 Jun 2025 11:01
 Operator : JC/MD
 Sample : Q2355-01
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 3897

Quant Time: Jun 24 04:02:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	104858	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	181206	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	162341	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	79106	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70070	48.312	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.620%		
35) Dibromofluoromethane	5.391	113	57671	48.103	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.200%		
50) Toluene-d8	8.647	98	217054	50.391	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	100.780%		
62) 4-Bromofluorobenzene	11.079	95	79651	49.606	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	99.220%		
Target Compounds						Qvalue
5) Bromomethane	1.612	94	1493	2.059	ug/l	91
10) Methyl Iodide	2.459	142	6581	8.334	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062325\
Data File : VX046809.D
Acq On : 23 Jun 2025 11:01
Operator : JC/MD
Sample : Q2355-01
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
3897

Quant Time: Jun 24 04:02:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

