

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062325\
 Data File : VX046811.D
 Acq On : 23 Jun 2025 11:43
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 24 04:03:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	171224	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	297676	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	270631	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	135211	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	112394	47.457	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.920%
35) Dibromofluoromethane	5.397	113	93869	47.661	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.320%
50) Toluene-d8	8.653	98	355662	50.263	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.520%
62) 4-Bromofluorobenzene	11.079	95	134279	50.908	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.820%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062325\
Data File : VX046811.D
Acq On : 23 Jun 2025 11:43
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 24 04:03:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046811.D
Acq: 23 Jun 2025 11:43




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:168 Resp: 171224
Ion Ratio Lower Upper
168 100
99 58.9 48.5 72.7



#33
1,2-Dichloroethane-d4
Concen: 47.457 ug/l
RT: 5.964 min Scan# 800
Delta R.T. -0.000 min
Lab File: VX046811.D
Acq: 23 Jun 2025 11:43

Tgt Ion: 65 Resp: 112394
Ion Ratio Lower Upper
65 100
67 52.5 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 911
Delta R.T. -0.000 min
Lab File: VX046811.D
Acq: 23 Jun 2025 11:43




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:114 Resp: 297676
Ion Ratio Lower Upper
114 100
63 20.4 0.0 44.2
88 16.4 0.0 33.2



#35
Dibromofluoromethane
Concen: 47.661 ug/l
RT: 5.397 min Scan# 707
Delta R.T. -0.000 min
Lab File: VX046811.D
Acq: 23 Jun 2025 11:43

Tgt Ion:113 Resp: 93869
Ion Ratio Lower Upper
113 100
111 103.3 82.0 123.0
192 19.3 15.3 22.9









#50
Toluene-d8
Concen: 50.263 ug/l
RT: 8.653 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046811.D
Acq: 23 Jun 2025 11:43




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 98 Resp: 355662
Ion Ratio Lower Upper
98 100
100 67.7 53.0 79.4



#62
4-Bromofluorobenzene
Concen: 50.908 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046811.D
Acq: 23 Jun 2025 11:43

Tgt Ion: 95 Resp: 134279
Ion Ratio Lower Upper
95 100
174 76.0 0.0 150.4
176 73.9 0.0 145.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046811.D
Acq: 23 Jun 2025 11:43

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Tgt Ion:117 Resp: 270631
Ion Ratio Lower Upper
117 100
82 54.3 44.6 66.8
119 31.8 25.8 38.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046811.D
Acq: 23 Jun 2025 11:43

Tgt Ion:152 Resp: 135211
Ion Ratio Lower Upper
152 100
115 61.3 43.2 129.6
150 157.2 0.0 346.8

