

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062419\
 Data File : VX010402.D
 Acq On : 24 Jun 2019 10:46
 Operator : JC/SP
 Sample : VX0624WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0624WBL01

Quant Time: Jun 25 03:48:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	272429	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	412924	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	359231	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	159389	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	134294	52.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.16%	
35) Dibromofluoromethane	5.49	113	134918	53.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.80%	
50) Toluene-d8	8.71	98	505906	52.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.68%	
62) 4-Bromofluorobenzene	11.13	95	166075	47.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.18%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062419\
Data File : VX010402.D
Acq On : 24 Jun 2019 10:46
Operator : JC/SP
Sample : VX0624WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0624WBL01

Quant Time: Jun 25 03:48:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

