

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX062419\
 Data File : VX010408.D
 Acq On : 24 Jun 2019 13:17
 Operator : JC/SP
 Sample : K3500-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-B-061919

Manual Integrations
 APPROVED

MMDadoda
 6/25/2019 9:01:41 AM

Quant Time: Jun 25 04:06:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	309556	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	471409	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	425556	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	203990	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	143073	49.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.58%	
35) Dibromofluoromethane	5.50	113	143652	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
50) Toluene-d8	8.71	98	551786	50.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.02%	
62) 4-Bromofluorobenzene	11.13	95	190531	47.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.66%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	4938	3.395	ug/l	93
30) Chloroform	5.22	83	7643	1.543	ug/l	87
31) Cyclohexane	5.58	56	17886	4.266	ug/l	91
39) Methylcyclohexane	7.46	83	29672	6.026	ug/l	100
64) Tetrachloroethene	9.34	164	4225	1.227	ug/l	94
67) Ethyl Benzene	10.25	91	18043	1.237	ug/l	94
68) m/p-Xylenes	10.35	106	4512	0.798	ug/l	99
73) Isopropylbenzene	11.02	105	210307	15.209	ug/l	98
78) n-propylbenzene	11.36	91	1376135	90.550	ug/l	100
80) 1,3,5-Trimethylbenzene	11.51	105	1477164	129.581	ug/l	100
83) tert-Butylbenzene	11.77	119	6205m	0.542	ug/l	
84) 1,2,4-Trimethylbenzene	11.80	105	4027957	351.695	ug/l	99
85) sec-Butylbenzene	11.94	105	260143	19.365	ug/l	96
86) p-Isopropyltoluene	12.06	119	156511	12.595	ug/l	98
89) n-Butylbenzene	12.38	91	955409	90.483	ug/l #	18
95) Naphthalene	13.83	128	61184	4.286	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062419\
Data File : VX010408.D
Acq On : 24 Jun 2019 13:17
Operator : JC/SP
Sample : K3500-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-B-061919

Manual Integrations
APPROVED

MMDadoda
6/25/2019 9:01:41 AM

Quant Time: Jun 25 04:06:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

