

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062419\
 Data File : VX010417.D
 Acq On : 24 Jun 2019 17:24
 Operator : JC/SP
 Sample : K3155-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 N43502

Quant Time: Jun 25 04:27:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	281353	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	431245	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	388896	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	178068	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	131134	49.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.42%	
35) Dibromofluoromethane	5.50	113	130521	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
50) Toluene-d8	8.71	98	504987	50.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.06%	
62) 4-Bromofluorobenzene	11.13	95	169915	46.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.24%	
Target Compounds						
16) Acetone	2.45	43	10429	7.888	ug/l	92
49) 1,4-Dioxane	7.74	88	3903	49.973	ug/l	94
51) 4-Methyl-2-Pentanone	8.64	43	4848	1.300	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062419\
Data File : VX010417.D
Acq On : 24 Jun 2019 17:24
Operator : JC/SP
Sample : K3155-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
N43502

Quant Time: Jun 25 04:27:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

