

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062419\
 Data File : VX010421.D
 Acq On : 24 Jun 2019 18:58
 Operator : JC/SP
 Sample : K3155-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BAT-B04

Quant Time: Jun 25 04:35:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	279422	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	424191	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	386009	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	178437	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	129782	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
35) Dibromofluoromethane	5.50	113	129619	49.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.88%	
50) Toluene-d8	8.71	98	505658	50.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.86%	
62) 4-Bromofluorobenzene	11.13	95	167979	46.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.72%	
Target Compounds						
3) Chloromethane	1.33	50	3416	1.510	ug/l	99
16) Acetone	2.45	43	7089	5.399	ug/l	95
44) Trichloroethene	7.23	130	2014	0.641	ug/l	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062419\
Data File : VX010421.D
Acq On : 24 Jun 2019 18:58
Operator : JC/SP
Sample : K3155-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BAT-B04

Quant Time: Jun 25 04:35:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

