

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062419\
 Data File : VX010422.D
 Acq On : 24 Jun 2019 19:21
 Operator : JC/SP
 Sample : K3155-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BAT-B06

Quant Time: Jun 25 04:36:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	263464	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	406018	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	367641	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	172150	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	125120	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
35) Dibromofluoromethane	5.50	113	125068	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
50) Toluene-d8	8.71	98	475444	50.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.06%	
62) 4-Bromofluorobenzene	11.13	95	161666	47.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.24%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.33	50	7098	3.328	ug/l	99
16) Acetone	2.44	43	9727	7.857	ug/l	95
52) Toluene	8.78	92	2517	0.380	ug/l	79
95) Naphthalene	13.83	128	4964	0.412	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062419\
Data File : VX010422.D
Acq On : 24 Jun 2019 19:21
Operator : JC/SP
Sample : K3155-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BAT-B06

Quant Time: Jun 25 04:36:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

