


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX062420\
 Data File : VX016939.D
 Acq On : 24 Jun 2020 09:50
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062420W.M
 Title : SW846 8260
 Last Update : Wed Jun 24 13:10:00 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.3	16778	PASS
75	95	30	60	53.2	41974	PASS
95	95	100	100	100.0	78861	PASS
96	95	5	9	6.7	5294	PASS
173	174	0.00	2	1.5	897	PASS
174	95	50	100	74.5	58757	PASS
175	174	5	9	7.9	4613	PASS
176	174	95	101	97.6	57373	PASS
177	176	5	9	6.0	3419	PASS