

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
 Data File : VX022882.D
 Acq On : 24 Jun 2021 20:54
 Operator : JC/MD
 Sample : M2783-19MS 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D4-20210618MS

Quant Time: Jun 25 05:40:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	66091	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	128914	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	117442	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49855	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	57727	48.13	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.26%		
35) Dibromofluoromethane	5.397	113	40021	48.89	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.78%		
50) Toluene-d8	8.653	98	159572	49.74	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.48%		
62) 4-Bromofluorobenzene	11.085	95	57340	47.89	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.78%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.173	85	46988	50.39	ug/l	96
3) Chloromethane	1.295	50	47921	46.80	ug/l	99
4) Vinyl Chloride	1.374	62	46143	48.84	ug/l	99
5) Bromomethane	1.599	94	24216	46.52	ug/l	95
6) Chloroethane	1.679	64	27783	47.58	ug/l	96
7) Trichlorofluoromethane	1.886	101	66550	48.49	ug/l	94
8) Diethyl Ether	2.142	74	24865	46.93	ug/l	83
9) 1,1,2-Trichlorotrifluo...	2.325	101	43209	50.28	ug/l	90
10) Methyl Iodide	2.453	142	45324	44.34	ug/l	95
11) Tert butyl alcohol	2.983	59	60244	241.00	ug/l	98
12) 1,1-Dichloroethene	2.319	96	41430	48.32	ug/l	94
13) Acrolein	2.239	56	32103	206.49	ug/l	98
14) Allyl chloride	2.666	41	78608	48.26	ug/l	93
15) Acrylonitrile	3.075	53	148027	259.04	ug/l	99
16) Acetone	2.392	43	124710	234.54	ug/l	96
17) Carbon Disulfide	2.514	76	98855	42.85	ug/l	98
18) Methyl Acetate	2.715	43	64453	49.25	ug/l	92
19) Methyl tert-butyl Ether	3.123	73	154140	50.13	ug/l	97
20) Methylene Chloride	2.794	84	50057	47.48	ug/l	# 87
21) trans-1,2-Dichloroethene	3.093	96	45743	48.09	ug/l	95
22) Diisopropyl ether	3.770	45	171007	50.47	ug/l	# 92
23) Vinyl Acetate	3.733	43	748187	263.22	ug/l	95
24) 1,1-Dichloroethane	3.617	63	90840	50.18	ug/l	97
25) 2-Butanone	4.568	43	209680	253.58	ug/l	93
26) 2,2-Dichloropropane	4.477	77	61314	44.63	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	54082	51.23	ug/l	97
28) Bromochloromethane	4.910	49	41360	48.50	ug/l	# 79
29) Tetrahydrofuran	5.019	42	139202	261.72	ug/l	87
30) Chloroform	5.105	83	90702	49.81	ug/l	97
31) Cyclohexane	5.477	56	81618	48.20	ug/l	96
32) 1,1,1-Trichloroethane	5.391	97	73289	50.10	ug/l	96
36) 1,1-Dichloropropene	5.702	75	65992	48.19	ug/l	97
37) Ethyl Acetate	4.727	43	83153	52.43	ug/l	# 93
38) Carbon Tetrachloride	5.684	117	60910	53.04	ug/l	95
39) Methylcyclohexane	7.391	83	79111	49.53	ug/l	98
40) Benzene	6.044	78	199477	50.52	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
 Data File : VX022882.D
 Acq On : 24 Jun 2021 20:54
 Operator : JC/MD
 Sample : M2783-19MS 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D4-20210618MS

Quant Time: Jun 25 05:40:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.934	41	45342	51.89	ug/l	90
42) 1,2-Dichloroethane	6.092	62	74915	51.57	ug/l	97
43) Isopropyl Acetate	6.348	43	132065	51.79	ug/l #	92
44) Trichloroethene	7.129	130	85339	91.14	ug/l	89
45) 1,2-Dichloropropane	7.440	63	52235	51.29	ug/l	98
46) Dibromomethane	7.586	93	34905	51.66	ug/l #	85
47) Bromodichloromethane	7.830	83	64363	47.06	ug/l	95
48) Methyl methacrylate	7.696	41	61635	52.22	ug/l	89
49) 1,4-Dioxane	7.665	88	26709	1051.31	ug/l	95
51) 4-Methyl-2-Pentanone	8.580	43	422664	266.32	ug/l	90
52) Toluene	8.726	92	120654	50.55	ug/l	98
53) t-1,3-Dichloropropene	8.982	75	71623	44.52	ug/l	98
54) cis-1,3-Dichloropropene	8.373	75	77366	44.93	ug/l	93
55) 1,1,2-Trichloroethane	9.159	97	49565	51.76	ug/l	100
56) Ethyl methacrylate	9.122	69	83946	47.61	ug/l #	83
57) 1,3-Dichloropropane	9.311	76	88384	50.77	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.165	63	69	5.29	ug/l #	45
59) 2-Hexanone	9.433	43	327575	272.06	ug/l	88
60) Dibromochloromethane	9.525	129	43470	46.61	ug/l	98
61) 1,2-Dibromoethane	9.616	107	50061	51.47	ug/l	98
64) Tetrachloroethene	9.275	164	42942	52.03	ug/l #	89
65) Chlorobenzene	10.086	112	124867	50.49	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	42738	53.47	ug/l	95
67) Ethyl Benzene	10.195	91	234060	51.39	ug/l	99
68) m/p-Xylenes	10.305	106	173049	102.32	ug/l	96
69) o-Xylene	10.647	106	84747	51.48	ug/l	96
70) Styrene	10.659	104	137825	50.44	ug/l	96
71) Bromoform	10.805	173	24975	44.11	ug/l #	100
73) Isopropylbenzene	10.964	105	219000	51.91	ug/l	99
74) N-amyl acetate	10.848	43	114821	48.30	ug/l	89
75) 1,1,2,2-Tetrachloroethane	11.213	83	75316	51.98	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	87347	66.33	ug/l	90
77) Bromobenzene	11.201	156	45195	51.52	ug/l	82
78) n-propylbenzene	11.305	91	265080	51.93	ug/l	96
79) 2-Chlorotoluene	11.366	91	158436	51.41	ug/l	95
80) 1,3,5-Trimethylbenzene	11.451	105	183858	53.08	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.024	75	21095	44.15	ug/l #	83
82) 4-Chlorotoluene	11.457	91	180142	50.58	ug/l	94
83) tert-Butylbenzene	11.719	119	169988	51.97	ug/l	88
84) 1,2,4-Trimethylbenzene	11.756	105	182790	52.43	ug/l	98
85) sec-Butylbenzene	11.896	105	231105	52.39	ug/l	97
86) p-Isopropyltoluene	12.012	119	181379	51.74	ug/l	96
87) 1,3-Dichlorobenzene	11.969	146	87177	51.51	ug/l	97
88) 1,4-Dichlorobenzene	12.043	146	86470	48.86	ug/l	96
89) n-Butylbenzene	12.335	91	170747	51.23	ug/l	96
90) Hexachloroethane	12.543	117	29742	46.18	ug/l	74
91) 1,2-Dichlorobenzene	12.335	146	84745	51.30	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	12.945	75	15836	46.57	ug/l	74
93) 1,2,4-Trichlorobenzene	13.591	180	46274	50.51	ug/l	97
94) Hexachlorobutadiene	13.725	225	17050	47.50	ug/l	98
95) Naphthalene	13.780	128	197691	48.70	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	48437	52.49	ug/l	97


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
Data File : VX022882.D
Acq On : 24 Jun 2021 20:54
Operator : JC/MD
Sample : M2783-19MS 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS

Quant Time: Jun 25 05:40:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.006 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:168 Resp: 66091
Ion Ratio Lower Upper
168 100
99 70.0 45.8 68.6#



#2
Dichlorodifluoromethane
Concen: 50.39 ug/l
RT: 1.173 min Scan# 14
Delta R.T. 0.001 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 85 Resp: 46988
Ion Ratio Lower Upper
85 100
87 30.6 16.4 49.2









#3
Chloromethane
Concen: 46.80 ug/l
RT: 1.295 min Scan# 34
Delta R.T. 0.001 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 50 Resp: 47921
Ion Ratio Lower Upper
50 100
52 32.3 26.4 39.6



#4
Vinyl Chloride
Concen: 48.84 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 62 Resp: 46143
Ion Ratio Lower Upper
62 100
64 32.5 25.7 38.5









#5
Bromomethane
Concen: 46.52 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 94 Resp: 24216
Ion Ratio Lower Upper
94 100
96 91.0 76.4 114.6



#6
Chloroethane
Concen: 47.58 ug/l
RT: 1.679 min Scan# 97
Delta R.T. 0.001 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 64 Resp: 27783
Ion Ratio Lower Upper
64 100
66 33.8 25.4 38.0









#7
 Trichlorofluoromethane
 Concen: 48.49 ug/l
 RT: 1.886 min Scan# 131
 Delta R.T. -0.000 min
 Lab File: VX022882.D
 Acq: 24 Jun 2021 20:54




Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D4-20210618MS


Tgt Ion:101 Resp: 66550
 Ion Ratio Lower Upper
 101 100
 103 61.8 53.2 79.8



#8
 Diethyl Ether
 Concen: 46.93 ug/l
 RT: 2.142 min Scan# 173
 Delta R.T. -0.000 min
 Lab File: VX022882.D
 Acq: 24 Jun 2021 20:54

Tgt Ion: 74 Resp: 24865
 Ion Ratio Lower Upper
 74 100
 45 105.3 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.28 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.006 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:101 Resp: 43209
Ion Ratio Lower Upper
101 100
85 46.2 34.9 52.3
151 65.0 61.4 92.2



#10
Methyl Iodide
Concen: 44.34 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion:142 Resp: 45324
Ion Ratio Lower Upper
142 100
127 45.9 33.3 49.9
141 14.1 11.4 17.0









#11

Tert butyl alcohol
Concen: 241.00 ug/l
RT: 2.983 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS




Tgt Ion: 59 Resp: 60244
Ion Ratio Lower Upper
59 100
57 10.5 7.8 11.8




#12

1,1-Dichloroethene
Concen: 48.32 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.001 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 96 Resp: 41430
Ion Ratio Lower Upper
96 100
61 165.9 124.6 187.0
98 62.0 50.6 76.0









#13
Acrolein
Concen: 206.49 ug/l
RT: 2.239 min Scan# 189
Delta R.T. -0.006 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 56 Resp: 32103
Ion Ratio Lower Upper
56 100
55 72.1 56.4 84.6



#14
Allyl chloride
Concen: 48.26 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 41 Resp: 78608
Ion Ratio Lower Upper
41 100
39 68.8 51.6 77.4
76 31.6 30.5 45.7









#15
Acrylonitrile
Concen: 259.04 ug/l
RT: 3.075 min Scan# 326
Delta R.T. 0.001 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:	53	Resp:	148027
Ion	Ratio	Lower	Upper
53	100		
52	82.4	65.3	97.9
51	35.8	28.2	42.2



#16
Acetone
Concen: 234.54 ug/l
RT: 2.392 min Scan# 214
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion:	43	Resp:	124710
Ion	Ratio	Lower	Upper
43	100		
58	32.7	27.8	41.8









#17
Carbon Disulfide
Concen: 42.85 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 76 Resp: 98855
Ion Ratio Lower Upper
76 100
78 8.5 7.2 10.8



#18
Methyl Acetate
Concen: 49.25 ug/l
RT: 2.715 min Scan# 267
Delta R.T. 0.006 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 43 Resp: 64453
Ion Ratio Lower Upper
43 100
74 23.4 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 50.13 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 73 Resp: 154140
Ion Ratio Lower Upper
73 100
57 22.8 17.1 25.7



#20
Methylene Chloride
Concen: 47.48 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 84 Resp: 50057
Ion Ratio Lower Upper
84 100
49 133.1 88.6 132.8#
51 38.9 27.0 40.6
86 62.5 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 48.09 ug/l
RT: 3.093 min Scan# 329
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 96 Resp: 45743
Ion Ratio Lower Upper
96 100
61 139.0 107.7 161.5
98 57.7 50.4 75.6



#22
Diisopropyl ether
Concen: 50.47 ug/l
RT: 3.770 min Scan# 440
Delta R.T. 0.001 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 45 Resp: 171007
Ion Ratio Lower Upper
45 100
43 78.3 58.0 87.0
87 24.7 25.4 38.0#
59 10.9 10.2 15.2









#23
 Vinyl Acetate
 Concen: 263.22 ug/l
 RT: 3.733 min Scan# 434
 Delta R.T. 0.000 min
 Lab File: VX022882.D
 Acq: 24 Jun 2021 20:54




Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D4-20210618MS


Tgt Ion: 43 Resp: 748187
 Ion Ratio Lower Upper
 43 100
 86 10.4 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 50.18 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: VX022882.D
 Acq: 24 Jun 2021 20:54

Tgt Ion: 63 Resp: 90840
 Ion Ratio Lower Upper
 63 100
 98 6.2 3.8 11.2
 100 4.0 2.5 7.5









#25
2-Butanone
Concen: 253.58 ug/l
RT: 4.568 min Scan# 571
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 43 Resp: 209680
Ion Ratio Lower Upper
43 100
72 25.6 23.7 35.5



#26
2,2-Dichloropropane
Concen: 44.63 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.006 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 77 Resp: 61314
Ion Ratio Lower Upper
77 100
97 22.2 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 51.23 ug/l

RT: 4.495 min Scan# 559





Delta R.T. 0.000 min

Lab File: VX022882.D

Acq: 24 Jun 2021 20:54

Tgt Ion: 96 Resp: 54082

Ion	Ratio	Lower	Upper
96	100		
61	146.9	0.0	285.6
98	62.4	0.0	129.6



#28

Bromochloromethane

Concen: 48.50 ug/l

RT: 4.910 min Scan# 627




Delta R.T. 0.007 min


Lab File: VX022882.D

Acq: 24 Jun 2021 20:54

Tgt Ion: 49 Resp: 41360

Ion	Ratio	Lower	Upper
49	100		
129	2.0	0.0	4.0
130	65.2	68.2	102.2#





#29

Tetrahydrofuran

Concen: 261.72 ug/l

RT: 5.019 min Scan# 645





Delta R.T. 0.000 min

Lab File: VX022882.D

Acq: 24 Jun 2021 20:54

Tgt Ion:	42	Resp:	139202
Ion	Ratio	Lower	Upper
42	100		
72	43.0	41.7	62.5
71	39.5	38.4	57.6

Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS



#30

Chloroform

Concen: 49.81 ug/l




RT: 5.105 min Scan# 659


Delta R.T. -0.000 min

Lab File: VX022882.D

Acq: 24 Jun 2021 20:54

Tgt Ion:	83	Resp:	90702
Ion	Ratio	Lower	Upper
83	100		
85	62.5	52.0	78.0









#31
Cyclohexane
Concen: 48.20 ug/l
RT: 5.477 min Scan# 720
Delta R.T. 0.001 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:	56	Resp:	81618
Ion	Ratio	Lower	Upper
56	100		
69	30.5	25.3	37.9
84	84.3	71.4	107.0



#32
1,1,1-Trichloroethane
Concen: 50.10 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion:	97	Resp:	73289
Ion	Ratio	Lower	Upper
97	100		
99	64.5	51.5	77.3
61	48.5	34.2	51.4





#33

1,2-Dichloroethane-d4

Concen: 48.13 ug/l

RT: 5.964 min Scan# 800

Delta R.T. 0.000 min

Lab File: VX022882.D

Acq: 24 Jun 2021 20:54

Tgt Ion: 65 Resp: 57727

Ion Ratio Lower Upper

65 100





67 50.9 0.0 105.4

Instrument :

MSVOA_X

ClientSampleId :

RE132D4-20210618MS



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.000 min

Lab File: VX022882.D

Acq: 24 Jun 2021 20:54




Tgt Ion: 114 Resp: 128914


Ion Ratio Lower Upper

114 100

63 20.7 0.0 40.8

88 15.8 0.0 32.8





#35

Dibromofluoromethane

Concen: 48.89 ug/l

RT: 5.397 min Scan# 707

Delta R.T. 0.000 min

Lab File: VX022882.D

Acq: 24 Jun 2021 20:54

Tgt Ion:113 Resp: 40021

Ion Ratio Lower Upper

113 100

111 104.5 81.9 122.9





192 15.6 16.7 25.1#

Instrument :

MSVOA_X

ClientSampleId :

RE132D4-20210618MS



#36

1,1-Dichloropropene

Concen: 48.19 ug/l

RT: 5.702 min Scan# 757

Delta R.T. 0.000 min

Lab File: VX022882.D

Acq: 24 Jun 2021 20:54




Tgt Ion: 75 Resp: 65992


Ion Ratio Lower Upper

75 100

110 33.6 18.1 54.4

77 30.8 25.3 37.9





#37

Ethyl Acetate

Concen: 52.43 ug/l

RT: 4.727 min Scan# 597





Delta R.T. -0.000 min

Lab File: VX022882.D

Acq: 24 Jun 2021 20:54

Tgt Ion:	43	Resp:	83153
Ion Ratio	100	Lower	Upper
43	100		
61	13.3	12.8	19.2
70	10.0	10.5	15.7#

Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS



#38

Carbon Tetrachloride

Concen: 53.04 ug/l




RT: 5.684 min Scan# 754


Delta R.T. -0.006 min

Lab File: VX022882.D

Acq: 24 Jun 2021 20:54

Tgt Ion:	117	Resp:	60910
Ion Ratio	100	Lower	Upper
117	100		
119	89.8	76.2	114.2
121	29.0	24.8	37.2









#39
Methylcyclohexane
Concen: 49.53 ug/l
RT: 7.391 min Scan# 1034
Delta R.T. 0.006 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 83 Resp: 79111
Ion Ratio Lower Upper
83 100
55 72.0 58.9 88.3
98 46.3 38.1 57.1



#40
Benzene
Concen: 50.52 ug/l
RT: 6.044 min Scan# 813
Delta R.T. -0.006 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 78 Resp: 199477
Ion Ratio Lower Upper
78 100
77 23.0 18.9 28.3









#41
Methacrylonitrile
Concen: 51.89 ug/l
RT: 4.934 min Scan# 631
Delta R.T. 0.006 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:	41	Resp:	45342
Ion	Ratio	Lower	Upper
41	100		
39	55.5	42.5	63.7
67	67.4	65.0	97.6
52	28.1	25.3	37.9



#42
1,2-Dichloroethane
Concen: 51.57 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.006 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion:	62	Resp:	74915
Ion	Ratio	Lower	Upper
62	100		
98	8.9	0.0	20.2









#43
Isopropyl Acetate
Concen: 51.79 ug/l
RT: 6.348 min Scan# 863
Delta R.T. -0.006 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 43 Resp: 132065
Ion Ratio Lower Upper
43 100
61 18.8 18.2 27.4
87 11.9 12.5 18.7#



#44
Trichloroethene
Concen: 91.14 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion:130 Resp: 85339
Ion Ratio Lower Upper
130 100
95 110.1 0.0 198.0





#45

1,2-Dichloropropane

Concen: 51.29 ug/l

RT: 7.440 min Scan# 1042

Delta R.T. 0.001 min

Lab File: VX022882.D





Acq: 24 Jun 2021 20:54

Tgt Ion: 63 Resp: 52235

Ion Ratio Lower Upper

63 100

65 30.7 25.7 38.5



#46

Dibromomethane

Concen: 51.66 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX022882.D

Acq: 24 Jun 2021 20:54




Tgt Ion: 93 Resp: 34905


Ion Ratio Lower Upper

93 100

95 82.4 66.8 100.2

174 76.5 83.1 124.7#









#47
Bromodichloromethane
Concen: 47.06 ug/l
RT: 7.830 min Scan# 1106
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:	83	Resp:	64363
Ion	Ratio	Lower	Upper
83	100		
85	60.0	51.4	77.0
127	7.5	6.8	10.2



#48
Methyl methacrylate
Concen: 52.22 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.006 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion:	41	Resp:	61635
Ion	Ratio	Lower	Upper
41	100		
69	80.9	76.0	114.0
39	60.4	44.6	67.0









#49
1,4-Dioxane
Concen: 1051.31 ug/l
RT: 7.665 min Scan# 1079
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:	88	Resp:	26709
Ion	Ratio	Lower	Upper
88	100		
43	34.7	23.2	34.8
58	71.1	55.0	82.4



#50
Toluene-d8
Concen: 49.74 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion:	98	Resp:	159572
Ion	Ratio	Lower	Upper
98	100		
100	66.2	51.9	77.9









#51
4-Methyl-2-Pentanone
Concen: 266.32 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 43 Resp: 422664
Ion Ratio Lower Upper
43 100
58 37.8 35.4 53.0



#52
Toluene
Concen: 50.55 ug/l
RT: 8.726 min Scan# 1253
Delta R.T. 0.006 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 92 Resp: 120654
Ion Ratio Lower Upper
92 100
91 174.3 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 44.52 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 75 Resp: 71623
Ion Ratio Lower Upper
75 100
77 32.2 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 44.93 ug/l
RT: 8.373 min Scan# 1195
Delta R.T. 0.001 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54


Tgt Ion: 75 Resp: 77366
Ion Ratio Lower Upper
75 100
77 31.4 25.7 38.5
39 51.4 35.1 52.7








#55
1,1,2-Trichloroethane
Concen: 51.76 ug/l
RT: 9.159 min Scan# 1324
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS







Tgt Ion:	97	Resp:	49565
Ion	Ratio	Lower	Upper
97	100		
83	85.9	69.3	103.9
85	55.5	44.3	66.5
99	62.8	50.3	75.5



#56
Ethyl methacrylate
Concen: 47.61 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion:	69	Resp:	83946
Ion	Ratio	Lower	Upper
69	100		
41	69.6	46.8	70.2
39	42.8	24.7	37.1#









#57

1,3-Dichloropropane
Concen: 50.77 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS




Tgt Ion: 76 Resp: 88384
Ion Ratio Lower Upper
76 100
78 32.1 26.2 39.2




#58

2-Chloroethyl Vinyl ether
Concen: 5.29 ug/l
RT: 8.165 min Scan# 1161
Delta R.T. -0.079 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 63 Resp: 69
Ion Ratio Lower Upper
63 100
106 0.0 23.5 35.3#









#59
2-Hexanone
Concen: 272.06 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 43 Resp: 327575
Ion Ratio Lower Upper
43 100
58 52.6 30.9 92.8



#60
Dibromochloromethane
Concen: 46.61 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. 0.001 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 129 Resp: 43470
Ion Ratio Lower Upper
129 100
127 78.8 38.4 115.2









#61
1,2-Dibromoethane
Concen: 51.47 ug/l
RT: 9.616 min Scan# 1399
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:107 Resp: 50061
Ion Ratio Lower Upper
107 100
109 92.5 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 47.89 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 95 Resp: 57340
Ion Ratio Lower Upper
95 100
174 61.7 0.0 154.6
176 58.2 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:117 Resp: 117442
Ion Ratio Lower Upper
117 100
82 60.5 47.4 71.2
119 31.1 26.6 39.8



#64
Tetrachloroethene
Concen: 52.03 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion:164 Resp: 42942
Ion Ratio Lower Upper
164 100
166 118.7 103.4 155.2
129 106.5 70.8 106.2#
131 96.8 72.5 108.7









#65
Chlorobenzene
Concen: 50.49 ug/l
RT: 10.086 min Scan# 1476
Delta R.T. 0.001 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:112 Resp: 124867
Ion Ratio Lower Upper
112 100
114 31.6 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 53.47 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion:131 Resp: 42738
Ion Ratio Lower Upper
131 100
133 89.7 47.7 143.1
119 67.4 32.1 96.5









#67
Ethyl Benzene
Concen: 51.39 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 91 Resp: 234060
Ion Ratio Lower Upper
91 100
106 30.3 24.6 37.0



#68
m/p-Xylenes
Concen: 102.32 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion:106 Resp: 173049
Ion Ratio Lower Upper
106 100
91 209.2 162.8 244.2









#69
o-Xylene
Concen: 51.48 ug/l
RT: 10.647 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:106 Resp: 84747
Ion Ratio Lower Upper
106 100
91 222.5 108.1 324.2



#70
Styrene
Concen: 50.44 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. 0.001 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion:104 Resp: 137825
Ion Ratio Lower Upper
104 100
78 54.7 40.0 60.0
103 54.5 43.4 65.0





#71

Bromoform

Concen: 44.11 ug/l





RT: 10.805 min Scan# 1594

Delta R.T. 0.000 min

Lab File: VX022882.D

Acq: 24 Jun 2021 20:54

Tgt Ion:	173	Resp:	24975
Ion Ratio	Lower	Upper	
173	100		
175	48.2	23.9	71.8
254	0.0	0.1	0.1#



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l




RT: 12.024 min Scan# 1794


Delta R.T. 0.000 min

Lab File: VX022882.D

Acq: 24 Jun 2021 20:54

Tgt Ion:	152	Resp:	49855
Ion Ratio	Lower	Upper	
152	100		
115	98.3	41.1	123.3
150	174.5	0.0	349.0









#73
Isopropylbenzene
Concen: 51.91 ug/l
RT: 10.964 min Scan# 1620
Delta R.T. 0.001 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:105 Resp: 219000
Ion Ratio Lower Upper
105 100
120 25.6 13.2 39.6



#74
N-ethyl acetate
Concen: 48.30 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. 0.001 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 43 Resp: 114821
Ion Ratio Lower Upper
43 100
70 39.7 39.0 58.6
55 22.6 21.5 32.3
61 22.2 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 51.98 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 83 Resp: 75316
Ion Ratio Lower Upper
83 100
131 8.5 5.1 15.2
85 64.2 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 66.33 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 75 Resp: 87347
Ion Ratio Lower Upper
75 100
77 32.0 19.1 57.5









#77
Bromobenzene
Concen: 51.52 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:156 Resp: 45195
Ion Ratio Lower Upper
156 100
77 200.8 82.8 248.6
158 94.3 49.1 147.4



#78
n-propylbenzene
Concen: 51.93 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 91 Resp: 265080
Ion Ratio Lower Upper
91 100
120 21.2 11.5 34.5









#79
2-Chlorotoluene
Concen: 51.41 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 91 Resp: 158436
Ion Ratio Lower Upper
91 100
126 30.4 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 53.08 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.006 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 105 Resp: 183858
Ion Ratio Lower Upper
105 100
120 47.1 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 44.15 ug/l
RT: 11.024 min Scan# 1630
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 75 Resp: 21095
Ion Ratio Lower Upper
75 100
53 116.2 74.5 111.7#
89 44.4 37.9 56.9



#82
4-Chlorotoluene
Concen: 50.58 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 91 Resp: 180142
Ion Ratio Lower Upper
91 100
126 26.5 14.9 44.5









#83
tert-Butylbenzene
Concen: 51.97 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:119 Resp: 169988
Ion Ratio Lower Upper
119 100
91 67.3 27.9 83.7
134 23.1 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 52.43 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion:105 Resp: 182790
Ion Ratio Lower Upper
105 100
120 43.6 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 52.39 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. 0.000 min
 Lab File: VX022882.D
 Acq: 24 Jun 2021 20:54




Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D4-20210618MS


Tgt Ion:105 Resp: 231105
 Ion Ratio Lower Upper
 105 100
 134 18.7 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 51.74 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX022882.D
 Acq: 24 Jun 2021 20:54

Tgt Ion:119 Resp: 181379
 Ion Ratio Lower Upper
 119 100
 134 25.4 12.9 38.7
 91 26.8 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 51.51 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.006 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:146 Resp: 87177
Ion Ratio Lower Upper
146 100
111 44.9 20.2 60.5
148 63.2 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 48.86 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. 0.001 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion:146 Resp: 86470
Ion Ratio Lower Upper
146 100
111 44.8 19.6 58.8
148 63.0 31.6 95.0









#89
n-Butylbenzene
Concen: 51.23 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion: 91 Resp: 170747
Ion Ratio Lower Upper
91 100
92 53.3 27.4 82.0
134 22.7 13.2 39.5



#90
Hexachloroethane
Concen: 46.18 ug/l
RT: 12.543 min Scan# 1879
Delta R.T. 0.001 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 117 Resp: 29742
Ion Ratio Lower Upper
117 100
201 57.4 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 51.30 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.006 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:146 Resp: 84745
Ion Ratio Lower Upper
146 100
111 48.4 20.8 62.2
148 62.7 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 46.57 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion: 75 Resp: 15836
Ion Ratio Lower Upper
75 100
155 68.3 45.3 135.9
157 85.1 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 50.51 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54




Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS


Tgt Ion:180 Resp: 46274
Ion Ratio Lower Upper
180 100
182 94.5 47.7 143.1
145 35.5 15.0 45.0



#94
Hexachlorobutadiene
Concen: 47.50 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion:225 Resp: 17050
Ion Ratio Lower Upper
225 100
223 61.2 31.1 93.2
227 65.4 31.6 95.0









#95
Naphthalene
Concen: 48.70 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618MS

Tgt Ion:128 Resp: 197691
Ion Ratio Lower Upper
128 100
127 12.4 10.3 15.5
129 10.6 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 52.49 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX022882.D
Acq: 24 Jun 2021 20:54

Tgt Ion:180 Resp: 48437
Ion Ratio Lower Upper
180 100
182 96.1 47.6 142.9
145 36.8 16.1 48.3

