

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
 Data File : VX022869.D
 Acq On : 24 Jun 2021 13:37
 Operator : JC/MD
 Sample : M2783-14 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP03-20210617

Quant Time: Jun 25 05:38:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53017	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	108733	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	93827	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	34512	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52381	54.442	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.880%
35) Dibromofluoromethane	5.397	113	35918	52.019	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.040%
50) Toluene-d8	8.653	98	144961	53.568	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.140%
62) 4-Bromofluorobenzene	11.085	95	46592	46.133	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.260%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.325	96	362	0.526	ug/l	# 84
16) Acetone	2.386	43	679	1.592	ug/l	# 84
44) Trichloroethene	7.129	130	34118	43.200	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
Data File : VX022869.D
Acq On : 24 Jun 2021 13:37
Operator : JC/MD
Sample : M2783-14 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP03-20210617

Quant Time: Jun 25 05:38:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

