

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
 Data File : VX022871.D
 Acq On : 24 Jun 2021 15:51
 Operator : JC/MD
 Sample : M2783-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D5-20210618

Quant Time: Jun 25 05:38:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	55071	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	110601	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100774	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	37971	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54020	54.051	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.100%	
35) Dibromofluoromethane	5.391	113	38422	54.705	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.420%	
50) Toluene-d8	8.653	98	151723	55.120	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 110.240%	
62) 4-Bromofluorobenzene	11.085	95	51322	49.958	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.920%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	1072	1.497	ug/l	# 87
12) 1,1-Dichloroethene	2.319	96	361	0.505	ug/l	# 82
44) Trichloroethene	7.129	130	63484	79.025	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
Data File : VX022871.D
Acq On : 24 Jun 2021 15:51
Operator : JC/MD
Sample : M2783-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D5-20210618

Quant Time: Jun 25 05:38:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

