

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
 Data File : VX022873.D
 Acq On : 24 Jun 2021 16:38
 Operator : JC/MD
 Sample : M2783-04DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D3-20210617DL

Quant Time: Jun 25 05:39:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	54944	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	112792	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100473	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	37509	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52876	53.029	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.060%	
35) Dibromofluoromethane	5.397	113	37273	52.039	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.080%	
50) Toluene-d8	8.653	98	149948	53.417	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.840%	
62) 4-Bromofluorobenzene	11.085	95	49410	47.163	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.320%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	22145	30.996	ug/l	90
44) Trichloroethene	7.129	130	13374	16.325	ug/l	87
64) Tetrachloroethene	9.281	164	2946	4.172	ug/l	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
Data File : VX022873.D
Acq On : 24 Jun 2021 16:38
Operator : JC/MD
Sample : M2783-04DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20210617DL

Quant Time: Jun 25 05:39:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

