

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
 Data File : VX022874.D
 Acq On : 24 Jun 2021 17:02
 Operator : JC/MD
 Sample : M2783-06DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D1-20210617DL

Quant Time: Jun 25 05:39:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	51182	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	106035	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	94098	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	34812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	51969	55.950	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.900%
35) Dibromofluoromethane	5.397	113	36961	54.891	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.780%
50) Toluene-d8	8.653	98	146130	55.374	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	110.740%
62) 4-Bromofluorobenzene	11.085	95	47713	48.445	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.880%
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	1155	1.735	ug/l	95
44) Trichloroethene	7.129	130	35878	46.584	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
Data File : VX022874.D
Acq On : 24 Jun 2021 17:02
Operator : JC/MD
Sample : M2783-06DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE114D1-20210617DL

Quant Time: Jun 25 05:39:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

