

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
 Data File : VX022880.D
 Acq On : 24 Jun 2021 20:07
 Operator : JC/MD
 Sample : M2803-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D3-20210621

Quant Time: Jun 25 05:40:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	50127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	104179	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	91882	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	33054	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	49387	54.289	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.580%	
35) Dibromofluoromethane	5.391	113	35119	53.085	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.180%	
50) Toluene-d8	8.653	98	134483	51.868	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.740%	
62) 4-Bromofluorobenzene	11.085	95	45311	46.826	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.660%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	290037	444.976	ug/l	90
12) 1,1-Dichloroethene	2.325	96	2343	3.603	ug/l	92
16) Acetone	2.386	43	1188	2.946	ug/l #	87
27) cis-1,2-Dichloroethene	4.495	96	968	1.209	ug/l	92
44) Trichloroethene	7.135	130	23559	31.134	ug/l	92
64) Tetrachloroethene	9.275	164	8173	12.658	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
Data File : VX022880.D
Acq On : 24 Jun 2021 20:07
Operator : JC/MD
Sample : M2803-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D3-20210621

Quant Time: Jun 25 05:40:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

