

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
 Data File : VX022894.D
 Acq On : 25 Jun 2021 02:23
 Operator : JC/MD
 Sample : I.BLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 I.BLK

Quant Time: Jun 25 05:43:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	49253	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	103959	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	89499	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	30848	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	45726	51.157	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.320%
35) Dibromofluoromethane	5.397	113	31586	47.846	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.700%
50) Toluene-d8	8.653	98	126306	48.818	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.640%
62) 4-Bromofluorobenzene	11.085	95	39860	41.280	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.560%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
Data File : VX022894.D
Acq On : 25 Jun 2021 02:23
Operator : JC/MD
Sample : I.BLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
I.BLK

Quant Time: Jun 25 05:43:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

