

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
 Data File : VX022897.D
 Acq On : 25 Jun 2021 03:33
 Operator : JC/MD
 Sample : M2838-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE109D3-20210622

Quant Time: Jun 25 05:43:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	52195	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	108957	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97807	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35668	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	48670	51.381	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.760%		
35) Dibromofluoromethane	5.391	113	33898	48.992	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.980%		
50) Toluene-d8	8.653	98	134363	49.549	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.100%		
62) 4-Bromofluorobenzene	11.085	95	44888	44.354	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 88.700%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	1616	2.381	ug/l	95
12) 1,1-Dichloroethene	2.319	96	345	0.509	ug/l #	61
16) Acetone	2.392	43	750	1.786	ug/l	98
27) cis-1,2-Dichloroethene	4.507	96	449	0.539	ug/l	83
44) Trichloroethene	7.135	130	57020	72.050	ug/l	95
64) Tetrachloroethene	9.281	164	671	0.976	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
Data File : VX022897.D
Acq On : 25 Jun 2021 03:33
Operator : JC/MD
Sample : M2838-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE109D3-20210622

Quant Time: Jun 25 05:43:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

