

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
 Data File : VX022898.D
 Acq On : 25 Jun 2021 03:57
 Operator : JC/MD
 Sample : M2838-11 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D2-20210622

Quant Time: Jun 25 05:43:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	49275	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	104092	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	94166	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36033	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50777	56.782	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.560%
35) Dibromofluoromethane	5.397	113	35182	53.225	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.440%
50) Toluene-d8	8.653	98	141455	54.603	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.200%
62) 4-Bromofluorobenzene	11.085	95	48076	49.725	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.440%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	638	0.996	ug/l	88
12) 1,1-Dichloroethene	2.319	96	623	0.975	ug/l #	75
27) cis-1,2-Dichloroethene	4.495	96	779	0.990	ug/l	70
44) Trichloroethene	7.135	130	338125	447.220	ug/l	93
64) Tetrachloroethene	9.275	164	315	0.476	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
Data File : VX022898.D
Acq On : 25 Jun 2021 03:57
Operator : JC/MD
Sample : M2838-11 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D2-20210622

Quant Time: Jun 25 05:43:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

