

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
 Data File : VX022899.D
 Acq On : 25 Jun 2021 04:20
 Operator : JC/MD
 Sample : M2838-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE126D1-20210622

Quant Time: Jun 25 05:43:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	50824	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	104294	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	93087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	34946	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	46350	50.252	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.500%
35) Dibromofluoromethane	5.397	113	31672	47.822	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.640%
50) Toluene-d8	8.653	98	129510	49.895	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.800%
62) 4-Bromofluorobenzene	11.085	95	42467	43.838	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.680%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	1281	3.133	ug/l	89
44) Trichloroethene	7.135	130	30529	40.301	ug/l	98
64) Tetrachloroethene	9.281	164	609	0.931	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
Data File : VX022899.D
Acq On : 25 Jun 2021 04:20
Operator : JC/MD
Sample : M2838-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE126D1-20210622

Quant Time: Jun 25 05:43:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

