

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
 Data File : VX022902.D
 Acq On : 25 Jun 2021 05:31
 Operator : JC/MD
 Sample : M2838-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D3-20210622

Quant Time: Jun 25 08:09:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	49314	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	104461	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	94008	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35459	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	45642	51.000	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.000%
35) Dibromofluoromethane	5.397	113	31844	48.005	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.000%
50) Toluene-d8	8.653	98	129256	49.718	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.440%
62) 4-Bromofluorobenzene	11.085	95	42876	44.190	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.380%
Target Compounds						
44) Trichloroethene	7.135	130	1391	1.833	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
Data File : VX022902.D
Acq On : 25 Jun 2021 05:31
Operator : JC/MD
Sample : M2838-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D3-20210622

Quant Time: Jun 25 08:09:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

