

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
 Data File : VX022886.D
 Acq On : 24 Jun 2021 22:51
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 25 05:41:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	50.000	51.171	-2.3	91	0.00
3 P	Chloromethane	50.000	50.530	-1.1	98	0.00
4 C	Vinyl Chloride	50.000	50.046	-0.1#	92	0.00
5 T	Bromomethane	50.000	51.926	-3.9	104	0.00
6 T	Chloroethane	50.000	51.648	-3.3	95	0.00
7 T	Trichlorofluoromethane	50.000	51.150	-2.3	95	0.00
8 T	Diethyl Ether	50.000	50.087	-0.2	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.924	0.2	93	0.00
10 T	Methyl Iodide	50.000	48.735	2.5	97	0.00
11 T	Tert butyl alcohol	250.000	255.415	-2.2	99	0.00
12 CM	1,1-Dichloroethene	50.000	51.217	-2.4#	96	0.00
13 T	Acrolein	250.000	215.126	13.9	85	0.00
14 T	Allyl chloride	50.000	49.969	0.1	94	0.00
15 T	Acrylonitrile	250.000	272.534	-9.0	99	0.00
16 T	Acetone	250.000	245.211	1.9	97	0.00
17 T	Carbon Disulfide	50.000	44.357	11.3	92	0.00
18 T	Methyl Acetate	50.000	52.134	-4.3	100	0.00
19 T	Methyl tert-butyl Ether	50.000	52.696	-5.4	97	0.00
20 T	Methylene Chloride	50.000	49.467	1.1	95	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.237	1.5	94	0.00
22 T	Diisopropyl ether	50.000	53.083	-6.2	96	0.00
23 T	Vinyl Acetate	250.000	277.239	-10.9	97	0.00
24 P	1,1-Dichloroethane	50.000	51.023	-2.0	94	0.00
25 T	2-Butanone	250.000	268.690	-7.5	99	0.00
26 T	2,2-Dichloropropane	50.000	44.144	11.7	80	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.309	-4.6	96	0.00
28 T	Bromochloromethane	50.000	50.452	-0.9	96	0.00
29 T	Tetrahydrofuran	250.000	277.537	-11.0	100	0.00
30 C	Chloroform	50.000	51.203	-2.4#	96	0.00
31 T	Cyclohexane	50.000	49.863	0.3	93	0.00
32 T	1,1,1-Trichloroethane	50.000	53.121	-6.2	96	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.144	1.7	99	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	95	0.00
35 S	Dibromofluoromethane	50.000	50.148	-0.3	98	0.00
36 T	1,1-Dichloropropene	50.000	48.939	2.1	94	0.00
37 T	Ethyl Acetate	50.000	54.166	-8.3	102	0.00
38 T	Carbon Tetrachloride	50.000	52.540	-5.1	94	0.00
39 T	Methylcyclohexane	50.000	49.967	0.1	93	0.00
40 TM	Benzene	50.000	51.797	-3.6	97	0.00
41 T	Methacrylonitrile	50.000	51.326	-2.7	99	0.00
42 TM	1,2-Dichloroethane	50.000	52.235	-4.5	95	0.00
43 T	Isopropyl Acetate	50.000	53.167	-6.3	98	0.00
44 TM	Trichloroethene	50.000	51.801	-3.6	98	0.00
45 C	1,2-Dichloropropane	50.000	53.009	-6.0#	98	0.00
46 T	Dibromomethane	50.000	51.863	-3.7	99	0.00
47 T	Bromodichloromethane	50.000	48.366	3.3	99	0.00
48 T	Methyl methacrylate	50.000	54.521	-9.0	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
 Data File : VX022886.D
 Acq On : 24 Jun 2021 22:51
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 25 05:41:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1075.134	-7.5	98	0.00
50 S	Toluene-d8	50.000	50.759	-1.5	99	0.00
51 T	4-Methyl-2-Pentanone	250.000	278.575	-11.4	99	0.00
52 CM	Toluene	50.000	52.789	-5.6#	97	0.00
53 T	t-1,3-Dichloropropene	50.000	45.877	8.2	94	0.00
54 T	cis-1,3-Dichloropropene	50.000	46.858	6.3	93	0.00
55 T	1,1,2-Trichloroethane	50.000	52.522	-5.0	96	0.00
56 T	Ethyl methacrylate	50.000	48.984	2.0	96	0.00
57 T	1,3-Dichloropropane	50.000	51.849	-3.7	95	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	253.235	-1.3	97	0.00
59 T	2-Hexanone	250.000	282.605	-13.0	101	0.00
60 T	Dibromochloromethane	50.000	47.977	4.0	97	0.00
61 T	1,2-Dibromoethane	50.000	52.843	-5.7	98	0.00
62 S	4-Bromofluorobenzene	50.000	49.639	0.7	98	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	97	0.00
64 T	Tetrachloroethene	50.000	51.219	-2.4	97	0.00
65 PM	Chlorobenzene	50.000	49.240	1.5	96	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	54.209	-8.4	96	0.00
67 C	Ethyl Benzene	50.000	51.860	-3.7#	96	0.00
68 T	m/p-Xylenes	100.000	103.017	-3.0	97	0.00
69 T	o-Xylene	50.000	51.125	-2.3	96	0.00
70 T	Styrene	50.000	52.522	-5.0	97	0.00
71 P	Bromoform	50.000	44.905	10.2	98	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	95	0.00
73 T	Isopropylbenzene	50.000	52.609	-5.2	95	0.00
74 T	N-ethyl acetate	50.000	50.464	-0.9	99	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	53.076	-6.2	96	0.00
76 T	1,2,3-Trichloropropane	50.000	54.087	-8.2	98	0.00
77 T	Bromobenzene	50.000	52.317	-4.6	98	0.00
78 T	n-propylbenzene	50.000	53.445	-6.9	95	0.00
79 T	2-Chlorotoluene	50.000	52.849	-5.7	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	54.051	-8.1	97	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.373	9.3	92	0.00
82 T	4-Chlorotoluene	50.000	52.105	-4.2	98	0.00
83 T	tert-Butylbenzene	50.000	52.389	-4.8	95	0.00
84 T	1,2,4-Trimethylbenzene	50.000	54.277	-8.6	95	0.00
85 T	sec-Butylbenzene	50.000	53.502	-7.0	95	0.00
86 T	p-Isopropyltoluene	50.000	52.969	-5.9	94	0.00
87 T	1,3-Dichlorobenzene	50.000	51.838	-3.7	95	0.00
88 T	1,4-Dichlorobenzene	50.000	50.443	-0.9	97	0.00
89 T	n-Butylbenzene	50.000	52.987	-6.0	93	0.00
90 T	Hexachloroethane	50.000	46.388	7.2	95	0.00
91 T	1,2-Dichlorobenzene	50.000	52.569	-5.1	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.099	3.8	101	0.00
93 T	1,2,4-Trichlorobenzene	50.000	53.115	-6.2	96	0.00
94 T	Hexachlorobutadiene	50.000	49.646	0.7	97	0.00
95 T	Naphthalene	50.000	50.708	-1.4	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
Data File : VX022886.D
Acq On : 24 Jun 2021 22:51
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 25 05:41:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	54.752	-9.5	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6