

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
 Data File : VX022866.D
 Acq On : 24 Jun 2021 12:15
 Operator : JC/MD
 Sample : VX0624WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0624WBL01

Quant Time: Jun 25 05:37:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	56595	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	116410	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	104153	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38983	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51001	49.656	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.320%
35) Dibromofluoromethane	5.397	113	36025	48.733	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.460%
50) Toluene-d8	8.653	98	143588	49.561	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.120%
62) 4-Bromofluorobenzene	11.085	95	47273	43.720	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.440%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062421\
Data File : VX022866.D
Acq On : 24 Jun 2021 12:15
Operator : JC/MD
Sample : VX0624WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0624WBL01

Quant Time: Jun 25 05:37:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.006 min
Lab File: VX022866.D
Acq: 24 Jun 2021 12:15




Instrument :
MSVOA_X
ClientSampleId :
VX0624WBL01


Tgt Ion:168 Resp: 56595
Ion Ratio Lower Upper
168 100
99 71.2 45.8 68.6#



#33
1,2-Dichloroethane-d4
Concen: 49.656 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX022866.D
Acq: 24 Jun 2021 12:15

Tgt Ion: 65 Resp: 51001
Ion Ratio Lower Upper
65 100
67 51.1 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX022866.D
Acq: 24 Jun 2021 12:15




Instrument :
MSVOA_X
ClientSampleId :
VX0624WBL01


Tgt Ion:114 Resp: 116410
Ion Ratio Lower Upper
114 100
63 22.2 0.0 40.8
88 14.6 0.0 32.8



#35
Dibromofluoromethane
Concen: 48.733 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX022866.D
Acq: 24 Jun 2021 12:15

Tgt Ion:113 Resp: 36025
Ion Ratio Lower Upper
113 100
111 102.8 81.9 122.9
192 15.5 16.7 25.1#









#50
Toluene-d8
Concen: 49.561 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX022866.D
Acq: 24 Jun 2021 12:15




Instrument :
MSVOA_X
ClientSampleId :
VX0624WBL01


Tgt Ion: 98 Resp: 143588
Ion Ratio Lower Upper
98 100
100 65.7 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 43.720 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022866.D
Acq: 24 Jun 2021 12:15

Tgt Ion: 95 Resp: 47273
Ion Ratio Lower Upper
95 100
174 62.5 0.0 154.6
176 59.9 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX022866.D
Acq: 24 Jun 2021 12:15

Instrument :
MSVOA_X
ClientSampleId :
VX0624WBL01

Tgt Ion:117 Resp: 104153
Ion Ratio Lower Upper
117 100
82 61.7 47.4 71.2
119 31.3 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX022866.D
Acq: 24 Jun 2021 12:15

Tgt Ion:152 Resp: 38983
Ion Ratio Lower Upper
152 100
115 65.5 41.1 123.3
150 154.6 0.0 349.0

