

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
 Data File : VX029762.D
 Acq On : 24 Jun 2022 09:55
 Operator : JC/MD
 Sample : VX0624WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0624WBL01

Quant Time: Jun 25 01:54:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	261849	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	438935	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	381282	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	168169	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	146616	42.299	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.600%
35) Dibromofluoromethane	5.385	113	128622	44.965	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.940%
50) Toluene-d8	8.653	98	491784	47.061	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.120%
62) 4-Bromofluorobenzene	11.079	95	178304	45.246	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
Data File : VX029762.D
Acq On : 24 Jun 2022 09:55
Operator : JC/MD
Sample : VX0624WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0624WBL01

Quant Time: Jun 25 01:54:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

