

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
 Data File : VX029769.D
 Acq On : 24 Jun 2022 12:49
 Operator : JC/MD
 Sample : N3366-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-2

Quant Time: Jun 25 01:59:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	262366	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	429907	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	381300	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	182297	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	136344	39.258	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	78.520%
35) Dibromofluoromethane	5.391	113	122568	43.749	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.500%
50) Toluene-d8	8.653	98	487036	47.586	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.180%
62) 4-Bromofluorobenzene	11.085	95	183016	47.417	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.840%
Target Compounds						
16) Acetone	2.379	43	5351	3.865	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
Data File : VX029769.D
Acq On : 24 Jun 2022 12:49
Operator : JC/MD
Sample : N3366-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-2

Quant Time: Jun 25 01:59:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

