

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
 Data File : VX029771.D
 Acq On : 24 Jun 2022 13:36
 Operator : JC/MD
 Sample : N3366-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-24

Quant Time: Jun 25 02:00:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	260112	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	433010	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	367076	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	171945	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	138490	40.222	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	80.440%
35) Dibromofluoromethane	5.391	113	126119	44.694	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.380%
50) Toluene-d8	8.653	98	481063	46.665	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.340%
62) 4-Bromofluorobenzene	11.085	95	181512	46.690	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.380%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.965	59	13212	8.676	ug/l	99
16) Acetone	2.380	43	1756	1.279	ug/l	88
22) Diisopropyl ether	3.770	45	17607	2.092	ug/l #	78
40) Benzene	6.044	78	49608	4.382	ug/l	94
83) tert-Butylbenzene	11.719	119	4563	0.431	ug/l	72
85) sec-Butylbenzene	11.896	105	3489	0.266	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
Data File : VX029771.D
Acq On : 24 Jun 2022 13:36
Operator : JC/MD
Sample : N3366-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-24

Quant Time: Jun 25 02:00:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

