

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
 Data File : VX029777.D
 Acq On : 24 Jun 2022 15:56
 Operator : JC/MD
 Sample : N3366-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-1

Quant Time: Jun 25 02:03:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	261659	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	438753	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	374383	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	167076	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	140692	40.620	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	81.240%
35) Dibromofluoromethane	5.391	113	127750	44.679	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.360%
50) Toluene-d8	8.653	98	485220	46.452	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.900%
62) 4-Bromofluorobenzene	11.085	95	177403	45.036	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.080%
Target Compounds						
16) Acetone	2.379	43	6239	4.519	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
Data File : VX029777.D
Acq On : 24 Jun 2022 15:56
Operator : JC/MD
Sample : N3366-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-1

Quant Time: Jun 25 02:03:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

