

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
 Data File : VX029792.D
 Acq On : 25 Jun 2022 02:20
 Operator : JC/MD
 Sample : N3386-01 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE105D2-20220613

Quant Time: Jun 25 05:32:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	269695	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	448044	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	389651	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172533	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	138237	38.722	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	77.440%		
35) Dibromofluoromethane	5.391	113	128077	43.864	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	87.720%		
50) Toluene-d8	8.653	98	495778	46.479	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	92.960%		
62) 4-Bromofluorobenzene	11.085	95	182634	45.403	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.800%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	2160	0.763	ug/l	92
44) Trichloroethene	7.129	130	227894	69.086	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
Data File : VX029792.D
Acq On : 25 Jun 2022 02:20
Operator : JC/MD
Sample : N3386-01 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE105D2-20220613

Quant Time: Jun 25 05:32:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

