

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
 Data File : VX029793.D
 Acq On : 25 Jun 2022 02:43
 Operator : JC/MD
 Sample : N3386-07 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D2-20220614

Quant Time: Jun 25 05:32:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	262463	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	436291	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	375894	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	164281	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	132460	38.126	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	76.260%
35) Dibromofluoromethane	5.391	113	124160	43.668	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.340%
50) Toluene-d8	8.653	98	487914	46.974	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.940%
62) 4-Bromofluorobenzene	11.079	95	173396	44.267	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.540%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.947	59	9954	4.043	ug/l #	75
12) 1,1-Dichloroethene	2.319	96	877	0.330	ug/l	94
44) Trichloroethene	7.129	130	459134	144.170	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
Data File : VX029793.D
Acq On : 25 Jun 2022 02:43
Operator : JC/MD
Sample : N3386-07 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D2-20220614

Quant Time: Jun 25 05:32:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

