

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
 Data File : VX029795.D
 Acq On : 25 Jun 2022 03:30
 Operator : JC/MD
 Sample : N3386-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE105D1-20220613

Quant Time: Jun 25 05:33:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	254920	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	419679	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	365149	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	161922	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	130147	38.568	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	77.140%
35) Dibromofluoromethane	5.391	113	122580	44.819	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.640%
50) Toluene-d8	8.653	98	467793	46.819	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.640%
62) 4-Bromofluorobenzene	11.079	95	167493	44.453	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.900%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	988	0.320	ug/l	82
9) 1,1,2-Trichlorotrifluo...	2.331	101	11245	4.204	ug/l	97
12) 1,1-Dichloroethene	2.319	96	1691	0.655	ug/l	88
27) cis-1,2-Dichloroethene	4.489	96	2890	0.876	ug/l	79
44) Trichloroethene	7.129	130	273628	88.882	ug/l	99
64) Tetrachloroethene	9.275	164	880	0.333	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
Data File : VX029795.D
Acq On : 25 Jun 2022 03:30
Operator : JC/MD
Sample : N3386-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE105D1-20220613

Quant Time: Jun 25 05:33:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

