

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
 Data File : VX029796.D
 Acq On : 25 Jun 2022 03:53
 Operator : JC/MD
 Sample : N3386-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Jun 25 05:33:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	259644	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	436489	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	377447	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	166596	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	134760	39.209	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	78.420%
35) Dibromofluoromethane	5.391	113	125499	44.119	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.240%
50) Toluene-d8	8.653	98	488272	46.987	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.980%
62) 4-Bromofluorobenzene	11.079	95	173238	44.207	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
Data File : VX029796.D
Acq On : 25 Jun 2022 03:53
Operator : JC/MD
Sample : N3386-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Jun 25 05:33:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

