

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
 Data File : VX029797.D
 Acq On : 25 Jun 2022 04:17
 Operator : JC/MD
 Sample : N3386-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW181S-20220614

Quant Time: Jun 25 05:33:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	249903	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	415481	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	358219	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	157749	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	127744	38.616	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	77.240%
35) Dibromofluoromethane	5.391	113	120734	44.590	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.180%
50) Toluene-d8	8.653	98	461114	46.617	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.240%
62) 4-Bromofluorobenzene	11.085	95	165561	44.384	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.760%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.325	96	1354	0.535	ug/l #	63
27) cis-1,2-Dichloroethene	4.495	96	2879	0.891	ug/l	81
30) Chloroform	5.099	83	2819	0.548	ug/l #	80
44) Trichloroethene	7.129	130	5226	0.581	ug/l	97
64) Tetrachloroethene	9.275	164	4102	1.583	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
Data File : VX029797.D
Acq On : 25 Jun 2022 04:17
Operator : JC/MD
Sample : N3386-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW181S-20220614

Quant Time: Jun 25 05:33:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

