

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
 Data File : VX029800.D
 Acq On : 25 Jun 2022 05:27
 Operator : JC/MD
 Sample : N3386-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D1-20220614

Quant Time: Jun 25 06:04:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	263915	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	432992	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	372520	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	163097	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	139095	39.815	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	79.640%
35) Dibromofluoromethane	5.391	113	126311	44.763	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.520%
50) Toluene-d8	8.653	98	486071	47.153	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.300%
62) 4-Bromofluorobenzene	11.079	95	172143	44.282	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.560%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	1217	0.439	ug/l	96
44) Trichloroethene	7.129	130	90674	27.763	ug/l	99
64) Tetrachloroethene	9.275	164	4735	1.758	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
Data File : VX029800.D
Acq On : 25 Jun 2022 05:27
Operator : JC/MD
Sample : N3386-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D1-20220614

Quant Time: Jun 25 06:04:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

