

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
 Data File : VX029802.D
 Acq On : 25 Jun 2022 06:13
 Operator : JC/MD
 Sample : N3386-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE117D1-20220615

Quant Time: Jun 25 09:05:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	257220	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	432540	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	372158	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	163785	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	140123	41.153	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	82.300%
35) Dibromofluoromethane	5.391	113	125000	44.345	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.700%
50) Toluene-d8	8.653	98	479248	46.540	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.080%
62) 4-Bromofluorobenzene	11.079	95	173052	44.562	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.120%
Target Compounds						
44) Trichloroethene	7.129	130	48257	14.251	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
Data File : VX029802.D
Acq On : 25 Jun 2022 06:13
Operator : JC/MD
Sample : N3386-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE117D1-20220615

Quant Time: Jun 25 09:05:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

