

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
 Data File : VX029803.D
 Acq On : 25 Jun 2022 06:36
 Operator : JC/MD
 Sample : N3386-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D1-20220616

Quant Time: Jun 25 09:05:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	262670	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	430259	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	373220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	163334	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	140168	40.313	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	80.620%
35) Dibromofluoromethane	5.391	113	127294	45.398	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.800%
50) Toluene-d8	8.653	98	484364	47.286	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.580%
62) 4-Bromofluorobenzene	11.085	95	172325	44.611	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.220%
Target Compounds						
					Qvalue	
30) Chloroform	5.092	83	1672	0.309	ug/l	84
44) Trichloroethene	7.129	130	10355	2.168	ug/l	96
64) Tetrachloroethene	9.275	164	1708	0.633	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062422\
Data File : VX029803.D
Acq On : 25 Jun 2022 06:36
Operator : JC/MD
Sample : N3386-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20220616

Quant Time: Jun 25 09:05:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

