

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
 Data File : VX036306.D
 Acq On : 24 Jun 2023 13:07
 Operator : JC/MD
 Sample : 03358-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-06-2023

Quant Time: Jun 26 03:24:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	232872	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	409871	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	363226	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	164968	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	168743	50.551	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.100%
35) Dibromofluoromethane	5.385	113	129108	47.772	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.540%
50) Toluene-d8	8.647	98	483781	47.927	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.860%
62) 4-Bromofluorobenzene	11.079	95	168272	47.327	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
Data File : VX036306.D
Acq On : 24 Jun 2023 13:07
Operator : JC/MD
Sample : 03358-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-06-2023

Quant Time: Jun 26 03:24:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

