

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
 Data File : VX036307.D
 Acq On : 24 Jun 2023 13:30
 Operator : JC/MD
 Sample : 03375-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-5-062123

Quant Time: Jun 26 03:24:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	185715	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	323340	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	291489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	133020	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	135454	50.882	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.760%
35) Dibromofluoromethane	5.391	113	101076	47.409	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.820%
50) Toluene-d8	8.647	98	388292	48.762	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.520%
62) 4-Bromofluorobenzene	11.079	95	135630	48.355	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
Data File : VX036307.D
Acq On : 24 Jun 2023 13:30
Operator : JC/MD
Sample : 03375-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-5-062123

Quant Time: Jun 26 03:24:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

