

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
 Data File : VX036308.D
 Acq On : 24 Jun 2023 13:54
 Operator : JC/MD
 Sample : 03319-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 26 03:24:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	215552	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	374810	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	337701	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	152995	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	146788	47.507	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.020%
35) Dibromofluoromethane	5.385	113	112539	45.537	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.080%
50) Toluene-d8	8.647	98	422874	45.812	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.620%#
62) 4-Bromofluorobenzene	11.079	95	146418	45.033	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.060%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
Data File : VX036308.D
Acq On : 24 Jun 2023 13:54
Operator : JC/MD
Sample : 03319-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 26 03:24:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

