

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
 Data File : VX036313.D
 Acq On : 24 Jun 2023 15:51
 Operator : JC/MD
 Sample : 03411-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 26 03:26:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	218765	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	394517	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	344393	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	150400	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	165820	52.879	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.760%
35) Dibromofluoromethane	5.385	113	126758	48.728	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.460%
50) Toluene-d8	8.647	98	476905	49.085	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.160%
62) 4-Bromofluorobenzene	11.079	95	159003	46.461	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
Data File : VX036313.D
Acq On : 24 Jun 2023 15:51
Operator : JC/MD
Sample : 03411-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 26 03:26:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

