

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
 Data File : VX036314.D
 Acq On : 24 Jun 2023 16:14
 Operator : JC/MD
 Sample : 03411-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jun 26 03:26:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	208061	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	367291	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	329444	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	145479	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	148602	49.826	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.660%
35) Dibromofluoromethane	5.385	113	111289	45.953	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.900%
50) Toluene-d8	8.646	98	419215	46.345	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.700%
62) 4-Bromofluorobenzene	11.079	95	143236	44.956	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.920%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
Data File : VX036314.D
Acq On : 24 Jun 2023 16:14
Operator : JC/MD
Sample : 03411-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jun 26 03:26:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

