

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
 Data File : VX036316.D
 Acq On : 24 Jun 2023 17:01
 Operator : JC/MD
 Sample : 03411-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Jun 26 03:27:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	191813	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	344564	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	314701	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139877	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	127519	46.379	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.760%
35) Dibromofluoromethane	5.385	113	97885	43.084	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.160%
50) Toluene-d8	8.647	98	371375	43.765	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	87.520%#
62) 4-Bromofluorobenzene	11.079	95	126782	42.417	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
Data File : VX036316.D
Acq On : 24 Jun 2023 17:01
Operator : JC/MD
Sample : 03411-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Jun 26 03:27:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

