

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
 Data File : VX036317.D
 Acq On : 24 Jun 2023 17:24
 Operator : JC/MD
 Sample : 03408-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-BR-05-390-410-062323

Quant Time: Jun 26 03:27:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	193352	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	340414	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	307611	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135825	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	138686	50.039	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.080%
35) Dibromofluoromethane	5.385	113	103924	46.300	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.600%
50) Toluene-d8	8.647	98	399119	47.607	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.220%
62) 4-Bromofluorobenzene	11.079	95	135716	45.959	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.920%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	1924	0.554	ug/l #	82
12) 1,1-Dichloroethene	2.325	96	231589	105.393	ug/l	92
16) Acetone	2.404	43	4862	4.104	ug/l	94
21) trans-1,2-Dichloroethene	3.093	96	19895	8.485	ug/l	99
24) 1,1-Dichloroethane	3.611	63	31859	7.555	ug/l	97
27) cis-1,2-Dichloroethene	4.489	96	153513	58.376	ug/l	86
29) Tetrahydrofuran	5.032	42	756	0.698	ug/l #	68
40) Benzene	6.044	78	4071	0.436	ug/l #	87
42) 1,2-Dichloroethane	6.092	62	20131	5.663	ug/l	100
44) Trichloroethene	7.129	130	158665	61.032	ug/l	96
49) 1,4-Dioxane	7.720	88	5473	91.017	ug/l #	93
52) Toluene	8.720	92	353543	59.687	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
Data File : VX036317.D
Acq On : 24 Jun 2023 17:24
Operator : JC/MD
Sample : 03408-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-BR-05-390-410-062323

Quant Time: Jun 26 03:27:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

