

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
 Data File : VX036301.D
 Acq On : 24 Jun 2023 11:04
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 26 03:20:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.473	0.488	-3.2	104	0.00
3 P	Chloromethane	0.821	0.763	7.1	91	0.00
4 C	Vinyl Chloride	0.898	0.861	4.1#	95	0.00
5 T	Bromomethane	0.657	0.627	4.6	98	0.00
6 T	Chloroethane	0.659	0.659	0.0	100	0.00
7 T	Trichlorofluoromethane	1.365	1.386	-1.5	102	0.00
8 T	Diethyl Ether	0.527	0.527	0.0	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.573	0.612	-6.8	104	0.00
10 T	Methyl Iodide	0.694	0.718	-3.5	97	0.00
11 T	Tert butyl alcohol	0.130	0.129	0.8	99	0.00
12 CM	1,1-Dichloroethene	0.568	0.566	0.4#	99	0.00
13 T	Acrolein	0.179	0.180	-0.6	107	0.00
14 T	Allyl chloride	0.923	0.985	-6.7	105	0.00
15 T	Acrylonitrile	0.331	0.336	-1.5	100	0.00
16 T	Acetone	0.306	0.330	-7.8	114	0.00
17 T	Carbon Disulfide	1.396	1.425	-2.1	103	0.00
18 T	Methyl Acetate	1.260	1.284	-1.9	102	0.00
19 T	Methyl tert-butyl Ether	1.836	1.910	-4.0	101	0.00
20 T	Methylene Chloride	0.660	0.633	4.1	99	0.00
21 T	trans-1,2-Dichloroethene	0.606	0.602	0.7	101	0.00
22 T	Diisopropyl ether	1.850	1.927	-4.2	102	0.00
23 T	Vinyl Acetate	1.343	1.445	-7.6	102	0.00
24 P	1,1-Dichloroethane	1.091	1.124	-3.0	101	0.00
25 T	2-Butanone	0.440	0.466	-5.9	105	0.00
26 T	2,2-Dichloropropane	0.773	0.919	-18.9	117	0.00
27 T	cis-1,2-Dichloroethene	0.680	0.698	-2.6	102	0.00
28 T	Bromochloromethane	0.566	0.485	14.3	91	0.00
29 T	Tetrahydrofuran	0.280	0.287	-2.5	102	0.00
30 C	Chloroform	1.098	1.126	-2.6#	102	0.00
31 T	Cyclohexane	0.918	0.964	-5.0	105	0.00
32 T	1,1,1-Trichloroethane	0.917	0.989	-7.9	105	0.00
33 S	1,2-Dichloroethane-d4	0.717	0.758	-5.7	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.330	0.361	-9.4	106	0.00
36 T	1,1-Dichloropropene	0.491	0.506	-3.1	102	0.00
37 T	Ethyl Acetate	0.501	0.533	-6.4	103	0.00
38 T	Carbon Tetrachloride	0.449	0.506	-12.7	107	0.00
39 T	Methylcyclohexane	0.589	0.653	-10.9	106	0.00
40 TM	Benzene	1.373	1.450	-5.6	102	0.00
41 T	Methacrylonitrile	0.278	0.294	-5.8	101	0.00
42 TM	1,2-Dichloroethane	0.522	0.554	-6.1	101	0.00
43 T	Isopropyl Acetate	0.799	0.868	-8.6	104	0.00
44 TM	Trichloroethene	0.382	0.392	-2.6	100	0.00
45 C	1,2-Dichloropropane	0.374	0.398	-6.4#	102	0.00
46 T	Dibromomethane	0.263	0.276	-4.9	102	0.00
47 T	Bromodichloromethane	0.476	0.524	-10.1	102	0.00
48 T	Methyl methacrylate	0.398	0.436	-9.5	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
 Data File : VX036301.D
 Acq On : 24 Jun 2023 11:04
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 26 03:20:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	87	-0.02
50 S	Toluene-d8	1.231	1.322	-7.4	102	0.00
51 T	4-Methyl-2-Pentanone	0.515	0.548	-6.4	101	0.00
52 CM	Toluene	0.870	0.897	-3.1#	99	0.00
53 T	t-1,3-Dichloropropene	0.481	0.573	-19.1	108	0.00
54 T	cis-1,3-Dichloropropene	0.543	0.621	-14.4	104	0.00
55 T	1,1,2-Trichloroethane	0.362	0.374	-3.3	99	0.00
56 T	Ethyl methacrylate	0.541	0.590	-9.1	102	0.00
57 T	1,3-Dichloropropane	0.620	0.638	-2.9	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.304	0.333	-9.5	99	0.00
59 T	2-Hexanone	0.376	0.418	-11.2	104	0.00
60 T	Dibromochloromethane	0.351	0.391	-11.4	103	0.00
61 T	1,2-Dibromoethane	0.384	0.402	-4.7	100	0.00
62 S	4-Bromofluorobenzene	0.434	0.487	-12.2	109	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.353	0.365	-3.4	99	0.00
65 PM	Chlorobenzene	1.051	1.113	-5.9	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.355	0.397	-11.8	103	0.00
67 C	Ethyl Benzene	1.869	2.002	-7.1#	102	0.00
68 T	m/p-Xylenes	0.726	0.772	-6.3	101	0.00
69 T	o-Xylene	0.701	0.752	-7.3	103	0.00
70 T	Styrene	1.148	1.227	-6.9	101	0.00
71 P	Bromoform	0.255	0.302	-18.4	107	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.768	3.988	-5.8	102	0.00
74 T	N-amyl acetate	1.490	1.734	-16.4	109	0.00
75 P	1,1,2,2-Tetrachloroethane	1.276	1.313	-2.9	101	0.00
76 T	1,2,3-Trichloropropane	1.067	1.106	-3.7	104	0.00
77 T	Bromobenzene	0.927	0.951	-2.6	102	0.00
78 T	n-propylbenzene	4.401	4.731	-7.5	104	0.00
79 T	2-Chlorotoluene	2.618	2.712	-3.6	102	0.00
80 T	1,3,5-Trimethylbenzene	3.191	3.402	-6.6	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.327	0.387	-18.3	114	0.00
82 T	4-Chlorotoluene	3.048	3.224	-5.8	103	0.00
83 T	tert-Butylbenzene	3.058	3.307	-8.1	105	0.00
84 T	1,2,4-Trimethylbenzene	3.157	3.360	-6.4	103	0.00
85 T	sec-Butylbenzene	3.874	4.207	-8.6	105	0.00
86 T	p-Isopropyltoluene	3.251	3.562	-9.6	105	0.00
87 T	1,3-Dichlorobenzene	1.703	1.814	-6.5	105	0.00
88 T	1,4-Dichlorobenzene	1.735	1.808	-4.2	105	0.00
89 T	n-Butylbenzene	2.894	3.275	-13.2	108	0.00
90 T	Hexachloroethane	0.492	0.564	-14.6	109	0.00
91 T	1,2-Dichlorobenzene	1.677	1.771	-5.6	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.242	0.278	-14.9	109	0.00
93 T	1,2,4-Trichlorobenzene	1.033	1.144	-10.7	109	0.00
94 T	Hexachlorobutadiene	0.423	0.470	-11.1	114	0.00
95 T	Naphthalene	3.413	3.717	-8.9	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062423\
Data File : VX036301.D
Acq On : 24 Jun 2023 11:04
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 26 03:20:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.008	1.108	-9.9	107	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6