

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062424\
 Data File : VX042016.D
 Acq On : 24 Jun 2024 19:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 25 06:16:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.381	0.347	8.9	85	0.00
3 P	Chloromethane	0.478	0.401	16.1	80	0.00
4 C	Vinyl Chloride	0.479	0.427	10.9#	84	0.00
5 T	Bromomethane	0.237	0.232	2.1	92	0.00
6 T	Chloroethane	0.241	0.252	-4.6	112	0.00
7 T	Trichlorofluoromethane	0.628	0.686	-9.2	106	0.00
8 T	Diethyl Ether	0.295	0.285	3.4	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.432	0.431	0.2	94	0.00
10 T	Methyl Iodide	0.537	0.607	-13.0	106	0.00
11 T	Tert butyl alcohol	0.109	0.107	1.8	89	-0.01
12 CM	1,1-Dichloroethene	0.431	0.418	3.0#	90	0.00
13 T	Acrolein	0.123	0.098	20.3	74	0.00
14 T	Allyl chloride	0.684	0.675	1.3	90	0.00
15 T	Acrylonitrile	0.286	0.301	-5.2	94	0.00
16 T	Acetone	0.253	0.245	3.2	97	0.00
17 T	Carbon Disulfide	0.879	0.715	18.7	80	0.00
18 T	Methyl Acetate	0.691	0.791	-14.5	99	0.00
19 T	Methyl tert-butyl Ether	1.425	1.502	-5.4	95	0.00
20 T	Methylene Chloride	0.554	0.509	8.1	93	0.00
21 T	trans-1,2-Dichloroethene	0.438	0.417	4.8	89	0.00
22 T	Diisopropyl ether	1.446	1.509	-4.4	94	0.00
23 T	Vinyl Acetate	1.235	1.283	-3.9	91	0.00
24 P	1,1-Dichloroethane	0.810	0.837	-3.3	95	0.00
25 T	2-Butanone	0.394	0.407	-3.3	92	0.00
26 T	2,2-Dichloropropane	0.586	0.554	5.5	87	0.00
27 T	cis-1,2-Dichloroethene	0.548	0.557	-1.6	92	0.00
28 T	Bromochloromethane	0.347	0.351	-1.2	97	0.00
29 T	Tetrahydrofuran	0.254	0.264	-3.9	92	0.00
30 C	Chloroform	0.843	0.889	-5.5#	97	0.00
31 T	Cyclohexane	0.674	0.628	6.8	89	0.00
32 T	1,1,1-Trichloroethane	0.705	0.740	-5.0	95	0.00
33 S	1,2-Dichloroethane-d4	0.577	0.541	6.2	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.355	0.344	3.1	92	0.00
36 T	1,1-Dichloropropene	0.377	0.364	3.4	91	0.00
37 T	Ethyl Acetate	0.491	0.516	-5.1	93	0.00
38 T	Carbon Tetrachloride	0.416	0.425	-2.2	93	0.00
39 T	Methylcyclohexane	0.476	0.450	5.5	89	0.00
40 TM	Benzene	1.271	1.273	-0.2	93	0.00
41 T	Methacrylonitrile	0.266	0.294	-10.5	96	0.00
42 TM	1,2-Dichloroethane	0.409	0.421	-2.9	94	0.00
43 T	Isopropyl Acetate	0.753	0.778	-3.3	92	0.00
44 TM	Trichloroethene	0.341	0.339	0.6	91	0.00
45 C	1,2-Dichloropropane	0.312	0.325	-4.2#	94	0.00
46 T	Dibromomethane	0.228	0.235	-3.1	93	0.00
47 T	Bromodichloromethane	0.418	0.449	-7.4	95	0.00
48 T	Methyl methacrylate	0.360	0.388	-7.8	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062424\
 Data File : VX042016.D
 Acq On : 24 Jun 2024 19:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 25 06:16:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	97	0.00
50 S	Toluene-d8	1.246	1.203	3.5	91	0.00
51 T	4-Methyl-2-Pentanone	0.519	0.556	-7.1	93	0.00
52 CM	Toluene	0.809	0.829	-2.5#	93	0.00
53 T	t-1,3-Dichloropropene	0.409	0.437	-6.8	92	0.00
54 T	cis-1,3-Dichloropropene	0.458	0.489	-6.8	92	0.00
55 T	1,1,2-Trichloroethane	0.337	0.360	-6.8	94	0.00
56 T	Ethyl methacrylate	0.524	0.570	-8.8	92	0.00
57 T	1,3-Dichloropropane	0.545	0.575	-5.5	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.244	0.262	-7.4	90	0.00
59 T	2-Hexanone	0.397	0.428	-7.8	92	0.00
60 T	Dibromochloromethane	0.366	0.404	-10.4	95	0.00
61 T	1,2-Dibromoethane	0.350	0.371	-6.0	95	0.00
62 S	4-Bromofluorobenzene	0.459	0.460	-0.2	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.364	0.350	3.8	92	0.00
65 PM	Chlorobenzene	1.040	1.061	-2.0	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.364	0.393	-8.0	95	0.00
67 C	Ethyl Benzene	1.630	1.709	-4.8#	94	0.00
68 T	m/p-Xylenes	0.658	0.681	-3.5	93	0.00
69 T	o-Xylene	0.651	0.687	-5.5	93	0.00
70 T	Styrene	1.092	1.185	-8.5	93	0.00
71 P	Bromoform	0.326	0.357	-9.5	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.040	3.234	-6.4	95	0.00
74 T	N-amyl acetate	1.350	1.385	-2.6	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.130	1.170	-3.5	95	0.00
76 T	1,2,3-Trichloropropane	0.908	0.926	-2.0	95	0.00
77 T	Bromobenzene	0.905	0.928	-2.5	93	0.00
78 T	n-propylbenzene	3.299	3.522	-6.8	94	0.00
79 T	2-Chlorotoluene	2.125	2.200	-3.5	95	0.00
80 T	1,3,5-Trimethylbenzene	2.532	2.674	-5.6	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.320	0.323	-0.9	88	0.00
82 T	4-Chlorotoluene	2.349	2.484	-5.7	94	0.00
83 T	tert-Butylbenzene	2.779	2.991	-7.6	96	0.00
84 T	1,2,4-Trimethylbenzene	2.541	2.739	-7.8	94	0.00
85 T	sec-Butylbenzene	3.143	3.408	-8.4	95	0.00
86 T	p-Isopropyltoluene	2.745	2.974	-8.3	94	0.00
87 T	1,3-Dichlorobenzene	1.592	1.644	-3.3	93	0.00
88 T	1,4-Dichlorobenzene	1.633	1.650	-1.0	93	0.00
89 T	n-Butylbenzene	2.113	2.284	-8.1	94	0.00
90 T	Hexachloroethane	0.467	0.497	-6.4	94	0.00
91 T	1,2-Dichlorobenzene	1.629	1.704	-4.6	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.224	0.244	-8.9	93	0.00
93 T	1,2,4-Trichlorobenzene	1.067	1.132	-6.1	93	0.00
94 T	Hexachlorobutadiene	0.539	0.566	-5.0	97	0.00
95 T	Naphthalene	3.289	3.572	-8.6	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062424\
Data File : VX042016.D
Acq On : 24 Jun 2024 19:47
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 25 06:16:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.125	1.183	-5.2	92	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6