

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062424\
 Data File : VX042017.D
 Acq On : 24 Jun 2024 20:11
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Jun 25 05:57:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 15 04:02:59 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	91	0.00
2 M	Dichlorodifluoromethane	1.942	1.680	13.5	78	0.00
3 M	Chloromethane	2.171	1.918	11.7	79	0.00
4 M	Vinyl Chloride	2.123	1.850	12.9	82	0.00
5 M	Bromomethane	1.135	0.953	16.0	80	0.00
6 M	Chloroethane	0.907	1.024	-12.9	103	0.00
7 M	Trichlorofluoromethane	2.674	2.355	11.9	79	0.00
8 T	Diethyl Ether	1.157	1.047	9.5	85	0.00
9	1,1,2-Trichlorotrifluoroeth	1.818	1.558	14.3	80	0.00
10 M	1,1-Dichloroethene	1.833	1.595	13.0	82	0.00
11	Methyl Iodide	2.386	2.257	5.4	87	0.00
12	Methyl Acetate	2.338	2.337	0.0	92	0.00
13 M	Acrolein	0.471	0.379	19.5	78	0.00
14 M	Acrylonitrile	1.109	1.033	6.9	85	0.00
15 M	Acetone	0.317	0.261	17.7	74	0.00
16 M	Carbon Disulfide	4.018	3.279	18.4	85	0.00
17	Allyl chloride	2.784	2.495	10.4	81	0.00
18 M	Methylene Chloride	2.103	1.928	8.3	83	0.00
19 M	trans-1,2-Dichloroethene	1.837	1.643	10.6	84	0.00
20 T	Diisopropyl ether	5.723	5.356	6.4	87	0.00
21 M	1,1-Dichloroethane	3.269	2.993	8.4	86	0.00
22 M	cis-1,2-Dichloroethene	2.223	2.013	9.4	85	0.00
23 M	tert-Butyl Alcohol	0.412	0.378	8.3	83	0.00
24 M	Methyl tert-Butyl Ether	5.671	5.284	6.8	87	0.00
25 M	Chloroform	3.322	3.130	5.8	87	0.00
26	Cyclohexane	2.840	2.346	17.4	78	0.00
27 s	1,2-Dichloroethane-d4	2.065	2.041	1.2	92	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
29	1,1-Dichloropropene	0.392	0.361	7.9	83	0.00
30 M	2-Butanone	0.262	0.257	1.9	83	0.00
31	2,2-Dichloropropane	0.429	0.355	17.2	75	0.00
32 M	1,1,1-Trichloroethane	0.496	0.479	3.4	87	0.00
33 M	Carbon Tetrachloride	0.428	0.408	4.7	86	0.00
34 M	Benzene	1.304	1.257	3.6	83	0.00
35	Methacrylonitrile	0.263	0.261	0.8	87	0.00
36 M	1,2-Dichloroethane	0.406	0.398	2.0	85	0.00
37 M	Trichloroethene	0.356	0.335	5.9	85	0.00
38	Methylcyclohexane	0.519	0.438	15.6	76	0.00
39 M	1,2-Dichloropropane	0.311	0.314	-1.0	89	0.00
40	Dibromomethane	0.230	0.223	3.0	86	0.00
41 M	Bromodichloromethane	0.412	0.406	1.5	89	0.00
42 M	Vinyl Acetate	0.876	0.853	2.6	85	0.00
43	Ethyl Acetate	0.491	0.486	1.0	88	0.00
44	Isopropyl Acetate	0.735	0.707	3.8	85	0.00
45 T	1,4-Dioxane	0.009	0.009	0.0	89	0.00
46	Methyl methacrylate	0.359	0.350	2.5	85	0.00
47	n-amyl Acetate	0.659	0.620	5.9	86	0.00
48 M	t-1,3-Dichloropropene	0.424	0.390	8.0	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062424\
 Data File : VX042017.D
 Acq On : 24 Jun 2024 20:11
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Jun 25 05:57:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 15 04:02:59 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.476	0.448	5.9	85	0.00
50 M	1,1,2-Trichloroethane	0.334	0.333	0.3	87	0.00
51	Ethyl methacrylate	0.531	0.502	5.5	86	0.00
52	1,3-Dichloropropane	0.546	0.538	1.5	86	0.00
53 M	Dibromochloromethane	0.366	0.353	3.6	89	0.00
54 M	1,2-Dibromoethane	0.352	0.352	0.0	86	0.00
55 M	2-Chloroethyl vinyl ether	0.247	0.243	1.6	83	0.00
56 M	Bromoform	0.298	0.279	6.4	89	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
58 M	4-Methyl-2-Pentanone	0.538	0.542	-0.7	86	0.00
59 M	2-Hexanone	0.415	0.413	0.5	85	0.00
60 S	4-Bromofluorobenzene	0.464	0.466	-0.4	88	0.00
61 M	Tetrachloroethene	0.377	0.349	7.4	79	0.00
62 M	Toluene	1.518	1.464	3.6	84	0.00
63 S	Toluene-d8	1.284	1.290	-0.5	86	0.00
64 M	Chlorobenzene	1.034	0.982	5.0	83	0.00
65	1,1,1,2-Tetrachloroethane	0.357	0.352	1.4	88	0.00
66 M	Ethyl Benzene	1.660	1.568	5.5	83	0.00
67 M	m/p-Xylenes	0.668	0.625	6.4	82	0.00
68 M	o-Xylene	0.666	0.628	5.7	83	0.00
69 M	Styrene	1.112	1.066	4.1	83	0.00
70	Isopropylbenzene	1.674	1.592	4.9	83	0.00
71 M	1,1,2,2-Tetrachloroethane	0.579	0.581	-0.3	86	0.00
72	1,2,3-Trichloropropane	0.469	0.460	1.9	86	0.00
73	Bromobenzene	0.488	0.468	4.1	85	0.00
74	n-propylbenzene	1.838	1.705	7.2	82	0.00
75	2-Chlorotoluene	1.167	1.114	4.5	83	0.00
76	1,3,5-Trimethylbenzene	1.402	1.324	5.6	82	0.00
77	t-1,4-Dichloro-2-butene	0.174	0.143	17.8	81	0.00
78	4-Chlorotoluene	1.297	1.212	6.6	83	0.00
79	tert-butylbenzene	1.488	1.407	5.4	83	0.00
80	1,2,4-Trimethylbenzene	1.401	1.335	4.7	84	0.00
81	sec-Butylbenzene	1.736	1.599	7.9	81	0.00
82	p-Isopropyltoluene	1.521	1.373	9.7	81	0.00
83 M	1,3-Dichlorobenzene	0.870	0.818	6.0	84	0.00
84 M	1,4-Dichlorobenzene	0.866	0.807	6.8	84	0.00
85	n-Butylbenzene	1.176	1.011	14.0	81	0.00
86 T	Hexachloroethane	0.248	0.221	10.9	84	0.00
87 M	1,2-Dichlorobenzene	0.880	0.820	6.8	83	0.00
88	1,2-Dibromo-3-Chloropropane	0.120	0.109	9.2	85	0.00
89	1,2,4-Trichlorobenzene	0.596	0.509	14.6	82	0.00
90	Hexachlorobutadiene	0.294	0.260	11.6	81	0.00
91 M	Naphthalene	1.800	1.662	7.7	85	0.00
92	1,2,3-Trichlorobenzene	0.617	0.568	7.9	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0