

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062424\
 Data File : VX042010.D
 Acq On : 24 Jun 2024 17:27
 Operator : JC/MD
 Sample : P3024-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 240620002-14-TRIP-BLANK

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/25/2024
 Supervised By : Mahesh Dadoda 06/25/2024

Quant Time: Jun 25 06:45:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 15 04:02:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	29446	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	154700	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	138967	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	60372	29.783	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.267%
60) 4-Bromofluorobenzene	11.079	95	56797	26.404	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	88.000%
63) Toluene-d8	8.646	98	179352	30.166	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.567%
Target Compounds						
25) Chloroform	5.092	83	6093m	1.868	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062424\
Data File : VX042010.D
Acq On : 24 Jun 2024 17:27
Operator : JC/MD
Sample : P3024-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
240620002-14-TRIP-BLANK

Quant Time: Jun 25 06:45:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 15 04:02:59 2024
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By : John Carlone 06/25/2024
Supervised By : Mahesh Dadoda 06/25/2024





#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 4.903 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX042010.D
Acq: 24 Jun 2024 17:27




Instrument :
MSVOA_X
ClientSampleId :
240620002-14-TRIP-BLANK



Tgt Ion:128 Resp: 2944
Ion Ratio Lower Upper
128 100
49 131.3 0.0 338.5
130 126.7 0.0 324.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/25/2024
Supervised By :Mahesh Dadoda 06/25/2024



#25
Chloroform
Concen: 1.868 ug/l m
RT: 5.092 min Scan# 657
Delta R.T. -0.000 min
Lab File: VX042010.D
Acq: 24 Jun 2024 17:27

Tgt Ion: 83 Resp: 6093
Ion Ratio Lower Upper
83 100
85 52.2 52.7 79.1#









#27
1,2-Dichloroethane-d4
Concen: 29.783 ug/l
RT: 5.958 min Scan# 798
Delta R.T. 0.006 min
Lab File: VX042010.D
Acq: 24 Jun 2024 17:27

Instrument :
MSVOA_X
ClientSampleId :
240620002-14-TRIP-BLANK

Tgt Ion: 65 Resp: 60372
Ion Ratio Lower Upper
65 100
67 54.2 43.4 65.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/25/2024
Supervised By :Mahesh Dadoda 06/25/2024



#28
1,4-Difluorobenzene
Concen: 30.000 ug/l
RT: 6.763 min Scan# 931
Delta R.T. -0.000 min
Lab File: VX042010.D
Acq: 24 Jun 2024 17:27


Tgt Ion:114 Resp: 154700
Ion Ratio Lower Upper
114 100
63 18.1 14.6 21.8
88 15.0 11.8 17.8





#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX042010.D
Acq: 24 Jun 2024 17:27




Instrument :
MSVOA_X
ClientSampleId :
240620002-14-TRIP-BLANK



Tgt Ion:117 Resp: 13896
Ion Ratio Lower Upper
117 100
82 51.2 41.1 61.7
119 32.6 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/25/2024
Supervised By :Mahesh Dadoda 06/25/2024



#60
4-Bromofluorobenzene
Concen: 26.404 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX042010.D
Acq: 24 Jun 2024 17:27


Tgt Ion: 95 Resp: 56797
Ion Ratio Lower Upper
95 100
174 96.1 78.9 118.3
176 92.8 77.1 115.7





#63
Toluene-d8
Concen: 30.166 ug/l
RT: 8.646 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX042010.D
Acq: 24 Jun 2024 17:27

Instrument :
MSVOA_X
ClientSampleId :
240620002-14-TRIP-BLANK



Tgt Ion: 98 Resp: 179352
Ion Ratio Lower Upper
98 100
100 66.1 52.7 79.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/25/2024
Supervised By :Mahesh Dadoda 06/25/2024

