

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062518\
 Data File : VX002867.D
 Acq On : 25 Jun 2018 19:01
 Operator : JC/MD
 Sample : VX0625WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0625WBL01

Quant Time: Jun 27 05:41:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	252342	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	350918	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	319672	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	180250	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	165658	47.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.00%	
35) Dibromofluoromethane	5.51	113	130478	46.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.90%	
50) Toluene-d8	8.72	98	492776	48.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.12%	
62) 4-Bromofluorobenzene	11.14	95	171985	43.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.88%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062518\
Data File : VX002867.D
Acq On : 25 Jun 2018 19:01
Operator : JC/MD
Sample : VX0625WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0625WBL01

Quant Time: Jun 27 05:41:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

