

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062518\
 Data File : VX002877.D
 Acq On : 25 Jun 2018 23:28
 Operator : JC/MD
 Sample : J3693-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-35

Quant Time: Jun 27 07:57:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	241766	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	337432	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	314021	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	180993	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	163448	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
35) Dibromofluoromethane	5.51	113	130293	48.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.52%	
50) Toluene-d8	8.72	98	476159	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
62) 4-Bromofluorobenzene	11.14	95	167936	44.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.26%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Methyl Iodide	2.51	142	2551	7.05	ug/l #	98
16) Acetone	2.45	43	2860	2.07	ug/l	96
19) Methyl tert-butyl Ether	3.21	73	19432	2.27	ug/l	100
22) Diisopropyl ether	3.88	45	4400	0.49	ug/l #	89
65) Chlorobenzene	10.14	112	19394	2.43	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062518\
Data File : VX002877.D
Acq On : 25 Jun 2018 23:28
Operator : JC/MD
Sample : J3693-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-35

Quant Time: Jun 27 07:57:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

