

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062518\
 Data File : VX002882.D
 Acq On : 26 Jun 2018 01:40
 Operator : JC/MD
 Sample : J3705-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-3

Quant Time: Jun 27 07:31:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	245999	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	346420	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	319410	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	179454	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	166309	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
35) Dibromofluoromethane	5.51	113	132956	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
50) Toluene-d8	8.72	98	490513	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
62) 4-Bromofluorobenzene	11.14	95	172346	44.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.22%	
Target Compounds						
10) Methyl Iodide	2.53	142	2385	6.98	ug/l	# 90
16) Acetone	2.45	43	6539	4.64	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062518\
Data File : VX002882.D
Acq On : 26 Jun 2018 01:40
Operator : JC/MD
Sample : J3705-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-3

Quant Time: Jun 27 07:31:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

